



**INTOXILYZER® 8000 INSTALLATION AND REPAIR CHECKOUT**

OFFICE OF ATTORNEY GENERAL  
CRIME LABORATORY DIVISION  
Toxicology Section/Breath Alcohol Program  
SFN 59281 (1-2010)

Serial Number: <u>80-004946</u>	Instrument Location: <u>Cass Jail</u>
Reason for Install/Repair:	
<input type="checkbox"/> Install After Lab Repair/Inspection <input checked="" type="checkbox"/> Other (Specify) <u>state tax sent this one out for install</u>	

Check When Done:

- A. Surge Protector Installed/Properly Grounded.
- B. Telephone Line Connected to Intoxilyzer® 8000.
- C. Breath Tube Heated.
- D. Review/Enter Preliminary Data Entry (Level 2, Function E).
- E. Review/Enter Gas Setup (Level1, Function S).
- F. Print Test (Level 1, Function P). (Attach Test Record.)
- G. ACA Test (Level 1, Function C). (Attach Test Record.)
- H. RFI Test (Use CMS Mode-First Room Air). (Attach Test Record.)
- I. Repair and/or Maintenance Performed: None, install from state tax.

[Signature]  
Field Inspector Signature

7-7-10  
Date

[Signature]  
Reviewed By

08 July 2010  
Date

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000            SN 80-004946  
Location = CASS                    8164.13.00 06/09  
07/07/2010                            10:45

\*\*\*\*\* Printer Test \*\*\*\*\*

abcdefghijklmnopqrstuvwxy1234567890-|=|  
ABCDEFGHIJKLMNopqrstuvwxyz!@#\$\$%^&\*()\_+?

abcdefghijklmnopqrstuvwxy1234567890-|=|  
ABCDEFGHIJKLMNopqrstuvwxyz!@#\$\$%^&\*()\_+?

Current Instrument Setup


Data Entry Mode:            Enabled  
Start Test Sequence:        DABACABA  
Display Prelim Rslt?        Yes  
Display Third Digit?        Yes  
Inhib Printer(Y/N)?        No  
Display Volume?            No  
Disable On Memfull?        No  
# of Print Copies?          1  
Select Std (D/W/I)?        Dry  
Standard Value?            0.080  
Standard Lot #?            709352  
Standard Cyl #?            14  
Standard Expiration?       12/23/2011  
Oper No?                    049524

Flow Cal. Date:            01/18/2010  
Slope                      717  
Intercept                  -728279

IR Calibration Date:       02/10/2010

	3um	9um
0th Coef(*100):	4321	-2310
1st Coef(*100):	254996	129019
2nd Coef(*100):	1560	1203
H2O adj(mg/l*10k):	410	499

\*\*\*\*\* Printer Test End \*\*\*\*\*

  
Operator Signature  
TROY A. HISCHER

Remarks:


Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004946  
Location = CASS      8164.13.00 06/09  
07/07/2010      10:46

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:47
02 Std. Gas	0.077	10:47
03 Room Air	0.000	10:47
04 Std. Gas	0.077	10:48
05 Room Air	0.000	10:48
06 Std. Gas	0.076	10:49
07 Room Air	0.000	10:49

Lot No = 709352  
Cyl No = 14  
Exp Date = 12/23/2011  
County = 09      Oper No. = 049524

  
Operator Signature  
TROY A. HISCHER

Remarks:

Form 106-I8000

80-004946  
p. 50 of 50  
JAM


Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004946  
Location = CASS      8164.13.00 06/09  
07/07/2010      11:04

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:05
02 Std. Gas	0.076	11:05
03 Room Air	0.000	11:06
04 Std. Gas	0.077	11:06
05 Room Air	0.000	11:07
06 Std. Gas	0.076	11:07
07 Room Air	0.000	11:08

Lot No = 709352  
Cyl No = 14  
Exp Date = 12/23/2011  
County = 09      Oper No. = 049524

  
Operator Signature  
TROY A. HISCHER

Remarks:

*2nd ACA*

Form 106-I8000

*80-004946  
P. 51 of 53  
YH*

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004946  
Location = CASS      8164.13.00 06/09  
07/07/2010      10:53

Test	AC	Time
01 Diagnostic	OK	11:03
02 Room Air	RFI*	11:03
03 Room Air	RFI*	11:03

\*Invalid Test  
Inhibited - RFI

Sub Name = INSTALLRFI, INSTALL I  
Sub DOB = 01/01/1980  
Sub Sex = Male      Weight = 200  
Test = DUI      Cit = RFI  
Dr. Lic. = ND/RFI  
Lot No = 709352  
Cyl No = 14  
Expiration Date = 12/23/2011  
County = 09      Oper No. = 049524

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature  
TROY A. HISCHER

Remarks:

Form 106-18000

80-004946  
P. 52 of 52  
AM