



ETHANOL BREATH STANDARD CYLINDER REPORT

OFFICE OF ATTORNEY GENERAL
CRIME LABORATORY DIVISION
Toxicology Section/Breath Alcohol Program
SFN 59282 (1-2010)

Only Chemical Test Operators may change the Ethanol Breath Standard and complete this form.

Chemical Test Operator Name: <u>Lt. Tracy L. Meidinger #258</u>		
Location: <u>UND-PD</u>		
Intoxilyzer® Serial Number: <u>80-004934</u>		
Gas Lot Number: <u>659352</u>	Gas Cylinder Number: <u>31</u>	Expiration Date: <u>09-30-2011</u>
<input type="checkbox"/> New Cylinder Installed	<input checked="" type="checkbox"/> Cylinder Not Changed	
Instrument Test Results (Report to 3 Digits; Example, 0.081%)		
Test 1: 0. <u>080</u> %		
Test 2: 0. <u>079</u> %		
Test 3: 0. <u>079</u> %		
(ATTACH TEST RECORD)		

Lt. Tracy L. Meidinger #258
Chemical Test Operator Signature

07-07-10
Date

Deb Kashur
Reviewed By

13 July 2010
Date

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004934
Location = UNDP 8164.13.00 06/09
07/07/2010 23:44

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	23:45
02 Std. Gas	0.080	23:45
03 Room Air	0.000	23:46
04 Std. Gas	0.079	23:46
05 Room Air	0.000	23:47
06 Std. Gas	0.079	23:47
07 Room Air	0.000	23:47

Lot No = 659358
Cyl No = 31
Exp Date = 09/03/2011
County = 18 Oper No. = 119409

Tracy L. Meidinger #258
Operator Signature
TRACY L. MEIDINGER

Remarks:

Form 106-I8000

ACA 07-07-10

80-004934
5/51 KM