



**ETHANOL BREATH STANDARD CYLINDER REPORT**  
OFFICE OF ATTORNEY GENERAL  
CRIME LABORATORY DIVISION  
Toxicology Section/Breath Alcohol Program  
SFN 59282 (1-2010)

Only Chemical Test Operators may change the Ethanol Breath Standard and complete this form.

Chemical Test Operator Name:	William R. Petty				
Location:	GF AF				
Intoxilyzer® Serial Number:	80-004949				
Gas Lot Number:	659358	Gas Cylinder Number:	39	Expiration Date:	09/03/2011
<input type="checkbox"/> New Cylinder Installed	<input checked="" type="checkbox"/> Cylinder Not Changed				
Instrument Test Results (Report to 3 Digits; Example, 0.081%)					
Test 1: 0.076 %					
Test 2: 0.077 %					
Test 3: 0.077 %					
(ATTACH TEST RECORD)					

William R. Petty  
Chemical Test Operator Signature

12 Jul 2010  
Date

Debo Keshur  
Reviewed By

22 Jul 2011  
Date

80-004949  
P. 94 of 95  
SM


Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004949  
Location = GFAP      8164.13.00 06/09  
07/12/2010      08:41

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	08:42
02 Std. Gas	0.076	08:42
03 Room Air	0.000	08:43
04 Std. Gas	0.077	08:43
05 Room Air	0.000	08:44
06 Std. Gas	0.077	08:44
07 Room Air	0.000	08:45

Lot No = 659358  
Cyl No = 39  
Exp Date = 09/03/2011  
County = 18      Oper No. = 120011



Operator Signature  
WILLIAM R. PETTY

Remarks:

Form 106-18000

80-004949  
P. 95 of 95  
PM