



ETHANOL BREATH STANDARD CYLINDER REPORT
 OFFICE OF ATTORNEY GENERAL
 CRIME LABORATORY DIVISION
 Toxicology Section/Breath Alcohol Program
 SFN 59282 (1-2010)

Only Chemical Test Operators may change the Ethanol Breath Standard and complete this form.

| | | |
|--|--------------------------------|--------------------------------|
| Chemical Test Operator Name: <u>Barry Vannatta</u> | | |
| Location: <u>Mohl</u> | | |
| Intoxilyzer® Serial Number: <u>80 004947</u> | | |
| Gas Lot Number: <u>659358</u> | Gas Cylinder Number: <u>27</u> | Expiration Date: <u>9-3-11</u> |
| <input type="checkbox"/> New Cylinder Installed <input checked="" type="checkbox"/> Cylinder Not Changed | | |
| Instrument Test Results (Report to 3 Digits; Example, 0.081%) Test 1: 0. <u>080</u> % Test 2: 0. <u>080</u> % Test 3: 0. <u>079</u> % | | |
| (ATTACH TEST RECORD) | | |

Barry Vannatta
 Chemical Test Operator Signature

7-15-11
 Date

Deb Kashner
 Reviewed By

20 July 2011
 Date


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004947
Location = MOHL 8164.13.00 06/09
07/15/2011 17:13

DRY CAL CHECK

| Test | AC | Time |
|-------------|-------|-------|
| 01 Room Air | 0.000 | 17:14 |
| 02 Std. Gas | 0.080 | 17:15 |
| 03 Room Air | 0.000 | 17:15 |
| 04 Std. Gas | 0.080 | 17:15 |
| 05 Room Air | 0.000 | 17:16 |
| 06 Std. Gas | 0.079 | 17:16 |
| 07 Room Air | 0.000 | 17:17 |

Lot No = 659358
Cyl No = 27
Exp Date = 09/03/2011
County = 38 Oper No. = 039909



Operator Signature
BARRY A. VANNATTA

Remarks:

Form 106-I8000

80-004947
P. 824 82
OM