

**ETHANOL BREATH STANDARD CYLINDER REPORT**

(Only Chemical Test Operators may change the Ethanol Breath Standard and complete this form.)

Chemical Test Operator Name: Phil Wolf

Location: Cavalier, ND

Intoxilyzer Serial No.: 80-002669

Lot Number: 659357

Cylinder Number: 9

Expiration Date: 9/3/11

Instrument Test Results (Report to 3 Digits; Example, 0.081%)

Test 1: 0.076 %

Test 2: 0.075 %

Test 3: 0.077 %

(ATTACH TEST RECORD)

P. Wolf  
Chemical Test Operator Signature

7/17/10  
Date

Deb Kashur  
Reviewed By

20 July 2010  
Date

80-00 2669  
p. 94 of 95  
PWW


Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000            SN 80-002669  
Location = CAVA                    8164.13.00 06/09  
07/17/2010                            15:03

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	15:04
02 Std. Gas	0.076	15:04
03 Room Air	0.000	15:05
04 Std. Gas	0.075	15:05
05 Room Air	0.000	15:06
06 Std. Gas	0.077	15:06
07 Room Air	0.000	15:06

Lot No = 659357  
Cyl No = 9  
Exp Date = 09/03/2011  
County = 34                            Oper No. = 120547

  
\_\_\_\_\_  
Operator Signature  
PHIL WOLF

Remarks:

Form 106-I8000

80-002669  
P. 95 of 95  
FM