



ETHANOL BREATH STANDARD CYLINDER REPORT
OFFICE OF ATTORNEY GENERAL
CRIME LABORATORY DIVISION
Toxicology Section/Breath Alcohol Program
SFN 59282 (1-2010)

Only Chemical Test Operators may change the Ethanol Breath Standard and complete this form.

Chemical Test Operator Name: <u>Dean Haaland</u>		
Location: <u>Cass County Jail</u>		
Intoxilyzer® Serial Number: <u>80-004946</u>		
Gas Lot Number: <u>709352</u>	Gas Cylinder Number: <u>14</u>	Expiration Date: <u>12-23-11</u>
<input type="checkbox"/> New Cylinder Installed	<input checked="" type="checkbox"/> Cylinder Not Changed	
Instrument Test Results (Report to 3 Digits; Example, 0.081%)		
Test 1: 0. <u>077</u> %		
Test 2: 0. <u>076</u> %		
Test 3: 0. <u>076</u> %		
(ATTACH TEST RECORD)		

Chemical Test Operator Signature

7-17-10
Date

Deb Kashner
Reviewed By

20 July 2010
Date


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004946
Location = CASS 8164.13.00 06/09
07/17/2010 14:00

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:01
02 Std. Gas	0.077	14:01
03 Room Air	0.000	14:02
04 Std. Gas	0.076	14:02
05 Room Air	0.000	14:03
06 Std. Gas	0.076	14:03
07 Room Air	0.000	14:04

Lot No = 709352
Cyl No = 14
Exp Date = 12/23/2011
County = 09 Oper No. = 109409



Operator Signature
DEAN J. HAALAND

Remarks:

Form 106-I8000

I hereby certify that the information in the
attached document is a true and correct copy
of the original document as reviewed by this office.



Deputy Sheriff
Cass County Sheriff's Office

80-004946
P. 54 of 54
JPM