

### BREATH ALCOHOL INSTRUMENT CHAIN OF CUSTODY WORKSHEET

Instrument: 8000 Serial Number: 80-004944

Transported Out of the Crime Laboratory:

Agency Being Assigned Instrument: \_\_\_\_\_

Name/Agency of Person Transporting Instrument:

\_\_\_\_\_ (Print Name) \_\_\_\_\_ (Print Agency)

Signature of Person Receiving Instrument: \_\_\_\_\_

Date Receiving Instrument: \_\_\_\_/\_\_\_\_/\_\_\_\_ From: \_\_\_\_\_

Or Shipped Via:

USPS \_\_\_\_ UPS \_\_\_\_ FedEx \_\_\_\_ Other \_\_\_\_\_

Date Sending Instrument: \_\_\_\_/\_\_\_\_/\_\_\_\_ Air Bill No. \_\_\_\_\_

Complete Air Bill (Identify Instrument/Serial Number) and attach a copy to this worksheet.

Transported to the Crime Laboratory:

Agency Returning Instrument: CMU

Name/Agency of Person Returning Instrument:

\_\_\_\_\_ (Print Name) \_\_\_\_\_ (Print Agency)

Signature of Person Returning Instrument: \_\_\_\_\_

Date Returning Instrument: 18 July 11 Received By: Debs Kashur

Or Received Via:

USPS \_\_\_\_ UPS \_\_\_\_ FedEx \_\_\_\_ Other \_\_\_\_\_

Date Receiving Instrument: \_\_\_\_/\_\_\_\_/\_\_\_\_ Air Bill No. \_\_\_\_\_

Attach shipping paperwork (if any) to this worksheet.

Instrument Transfer Information Entered Into the Equipment Database:

Date of Action 18 July 2011 By: Debs Kashur

80-004944  
P. 67 of 69  
km



# SERVICE WORK ORDER

DATE	5-11-11
RCVD VIA	FEDEX

316 E. 9<sup>th</sup> St., Owensboro KY 42303  
 Phone 866-835-0690  
 Fax: 270-685-6268

326608
--------

<b>BILL TO:</b>		
ATTN:		
CITY STATE ZIP		
CUSTOMER NO. 585013		
PHONE		

<b>SHIP TO:</b> Ofc. of Atty. Gen. Crime Lab		
ATTN: Deb Kashner		
2641 E. Main Ave.		
CITY	STATE	ZIP
Bismarck	ND	58501
PHONE 701-328-6159		
FAX 701-328-6185		

MODEL # 8000	SERIAL # 80-004944	BILL CODE EW
EXTRA PARTS RCVD BT, sensor, battery		
DESCRIPTION OF PROBLEM EEPROM FAIL.		

<b>Infrared / Fuel Cell</b> PRELIMINARY TEST INFORMATION	BATTERY- EXT <u>Disc</u> V, INT <u>3.02</u> V
FEEPROM FAIL	VERIFIED PROBLEM? YES
WORK PERFORMED: Replaced EEPROM. Battery monitor board Installed cap Rev CM11-076	
SEAL #04046	
<b>Miscellaneous</b> PRELIMINARY TEST INFORMATION	
WORK PERFORMED: RA	

PARTS USED		
QTY	PART NO.	DESCRIPTION
1	220286	EEPROM
1	BATT BOARD	0021509
1	110211	Cap.

REPAIR	CAL. ADJ.	CALIBRATOR	FINAL
LH.	DR	DR	DR
1/2 HOURS	HOURS	HOURS	HOURS
20	1	5	5
DATE	DATE	DATE	DATE
11 July 2011	7-1-11	7-7-11	7-7-11

SHIPPING STICKER HERE
SPECIAL SHIPPING INSTRUCTIONS
80-004944
P. 68 of 69
VM

# Certificate of Calibration

## **Intoxilyzer**® 8000

This is to certify that the calibration of **Intoxilyzer**® serial number 80-004944, manufactured by CMI Inc., a subsidiary of MPD, Inc. of Owensboro, Kentucky, was tested and found to be in compliance with the National Highway Traffic Safety Administration Standard for Devices to Measure Breath Alcohol (F.R., No. 179 48705-48710 Sept. 17, 1993) and with International Recommendation OIML R126 Edition 1998 (E). Calibration solutions are traceable to NIST standard material 1828.

Date July 7, 2011 Signed   
Technician

**CMI** INC.

80-004944  
P. 29 of 29  
CMI

316 East 9<sup>th</sup> Street  
Owensboro, KY 42303  
USA

Part No. 650519 CMI 7/9/09