

INTOXILYZER 8000 INSPECTION

Serial No.: 80-004944

Location: TOXL

Check When Done

- A. Is the warm-up time less than 20 minutes? ✓
- B. Is a three-pronged grounded outlet used? ✓
- C. Is the breath tube heated? ✓
- D. Is the diagnostic check complete? ✓
- E. Is the time, date, and year correct (re-set if necessary)? ✓
- F. Print test. (Attach test record.) ✓
- G. Low AC. Use <0.03 AC in ACA mode. (Attach test record.)
Sim SN DR5132 Lot No. Rep AC 0.021 ✓
- H. Linearity Test. Use ≥ 0.25 AC in ACA mode. (Attach test record.)
Sim SN DR5185 Lot No. Rep AC 0.297 ✓
- I. Interferent check. Use 0.05% acetone plus 0.10 AC ethanol in ABA mode.
(Attach test record.)
Sim SN DK3846 Lot No. Rep AC 0.10 AC + acetone ✓
- J. RFI check. Run CMS mode. Key radio on first room air.
(Attach test record.) ✓
- K. Calibration Check. Use Ethanol Breath Standard cylinder. Do three sets in ACA mode.
(Attach test record.)

Lot No. 05911080A1 Cyl No. 36 Exp. Date 02/01/2013
Test 1 0.079 Test 4 0.080 Test 7 0.080
Test 2 0.079 Test 5 0.079 Test 8 0.080
Test 3 0.079 Test 6 0.079 Test 9 0.079
Average 0.079

L. Tank pressure, Level 3 Func D Sub G P: 145 psi Reg 145 psi

M. Remarks/Maintenance record: Annual inspection after Eeprom change from CMI.

20 July 2011
Date

Deb Kashner
Field Inspector's Signature

7-20-11
Date

[Signature]
Reviewed by

80-004944
P. 87495
JFM

Intoxilyzer Test Record and Checklist
 NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
 North Dakota Model 8000 SN 80-004944
 Location = TOXL 8164.13.00 06/09
 07/18/2011 17:10

***** Printer Test *****

abcdefghijklmnopqrstuvwxyz1234567890-_=|
 ABCDEFGHIJKLMNOPQRSTUVWXYZ!@#\$\$%^&*()_+?

abcdefghijklmnopqrstuvwxyz1234567890-_=|
 ABCDEFGHIJKLMNOPQRSTUVWXYZ!@#\$\$%^&*()_+?

Current Instrument Setup

Data Entry Mode: Enabled
 Start Test Sequence: DABACABA
 Display Prelim Rslt? Yes
 Display Third Digit? Yes
 Inhib Printer(Y/N)? No
 Display Volume? No
 Disable On Memfull? Yes
 # of Print Copies? 1
 Select Std (D/W/I)? Dry
 Standard Value? 0.080
 Standard Lot #? 05911080A1
 Standard Cyl #? 36
 Standard Expiration? 02/01/2013
 Oper No? 777777


Flow Cal. Date: 07/01/2011
 Slope 685
 Intercept -634740

IR Calibration Date: 07/01/2011

	3um	9um

0th Coef(*100):	-22591	-17234
1st Coef(*100):	265915	132691
2nd Coef(*100):	2841	1562
H2O adj(mg/l*10k):	559	425

***** Printer Test End *****



 Operator Signature
 DEB KASHUR

Remarks:

Print test

*80-004944
 p. 88 of 95
 SEM*

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004944
Location = TOXL 8164.13.00 06/09
07/18/2011 16:55

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	16:56
02 Std. Sol.	0.019	16:57
03 Room Air	0.000	16:58
04 Std. Sol.	0.019	16:58
05 Room Air	0.000	16:59
06 Std. Sol.	0.020	17:00
07 Room Air	0.000	17:00

08 Sim Temp = 34.0°C

Simul Ser No = 0

Std Sol No = 0

County = 08

Oper No. = 777777

Deb Kasher

Operator Signature
DEB KASHUR

Remarks:

low AC 0.021AC

Form 106-I8000

80-004944
p. 89 of 95
km

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004944
Location = TOXL 8164.13.00 06/09
07/18/2011 17:01

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	17:02
02 Std. Sol.	0.297	17:02
03 Room Air	0.000	17:03
04 Std. Sol.	0.300	17:04
05 Room Air	0.000	17:04
06 Std. Sol.	0.300	17:05
07 Room Air	0.000	17:05


08 Sim Temp = 34.0°C

Simul Ser No = 0

Std Sol No = 0

County = 08

Oper No. = 777777


Operator Signature
DEB KASHUR

Remarks:

Linearity test 0.297 AC

Form 106-I8000

80-004944
P. 90 of 95
KM

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004944
Location = TOXL 8164.13.00 06/09
07/20/2011 07:59

Test	AC	Time
01 Room Air	0.000	08:00
02 *Subject Test	INT*	08:00
03 Room Air	0.000	08:01

*Invalid Test
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 02/01/1992
Sub Sex = Female Weight = 150
Test = DUI Cit = NA
Dr. Lic. = ND/DIS921456
Lot No = 05911080A1
Cyl No = 36
Expiration Date = 02/01/2013
County = 08 Oper No. = 777777

Deb Kashur

Operator Signature
DEB KASHUR

Remarks:

Interferent check 0.10AC + acetone

Form 106-I8000

80-004944
8.91.95
VFM

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004944
Location = TOXL 8164.13.00 06/09
07/20/2011 07:57

Test	AC	Time
01 Diagnostic	OK	07:58
02 Room Air	RFI*	07:58
03 Room Air	0.000	07:59

*Invalid Test
Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 02/01/1992
Sub Sex = Female Weight = 150
Test = DUI Cit = NA
Dr. Lic. = ND/DIS921456
Lot No = 05911080A1
Cyl No = 36
Expiration Date = 02/01/2013
County = 08 Oper No. = 777777

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.

Deb Kasher

Operator Signature
DEB KASHUR

Remarks:

RFI check

Form 106-I8000

80-004944
P-92495
VEM

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004944
Location = TOXL 8164.13.00 06/09
07/18/2011 17:06

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	17:07
02 Std. Gas	0.079	17:07
03 Room Air	0.000	17:08
04 Std. Gas	0.079	17:08
05 Room Air	0.000	17:09
06 Std. Gas	0.079	17:09
07 Room Air	0.000	17:10

Lot No = 05911080A1
Cyl No = 36
Exp Date = 02/01/2013
County = 08 Oper No. = 777777

Deb Kasher

Operator Signature
DEB KASHUR

Remarks:

Calibration check 0.080 AC

Form 106-I8000

80-004944
P. 93 of 95
EM

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004944
Location = TOXL 8164.13.00 06/09
07/18/2011 17:10

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	17:11
02 Std. Gas	0.080	17:11
03 Room Air	0.000	17:12
04 Std. Gas	0.079	17:12
05 Room Air	0.000	17:12
06 Std. Gas	0.079	17:13
07 Room Air	0.000	17:13

Lot No = 05911080A1
Cyl No = 36
Exp Date = 02/01/2013
County = 08 Oper No. = 777777

Deb Kashur
Operator Signature
DEB KASHUR

Remarks:

Calibration check 0.080 AC

Form 106-I8000

80-004944
P.94-195
JTM

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004944
Location = TOXL 8164.13.00 06/09
07/18/2011 17:14

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	17:14
02 Std. Gas	0.080	17:14
03 Room Air	0.000	17:15
04 Std. Gas	0.080	17:15
05 Room Air	0.000	17:16
06 Std. Gas	0.079	17:16
07 Room Air	0.000	17:17

Lot No = 05911080A1
Cyl No = 36
Exp Date = 02/01/2013
County = 08 Oper No. = 777777

Deb Kashur
Operator Signature
DEB KASHUR

Remarks:

Calibration check 0.080 AC

Form 106-I8000

80-004944
p. 95 of 95
KAM