

BREATH ALCOHOL INSTRUMENT CHAIN OF CUSTODY WORKSHEET

Instrument: Intox 8000 Serial Number: 80-004204

Transported Out of the Crime Laboratory:

Agency Being Assigned Instrument: _____

Name/Agency of Person Transporting Instrument:

(Print Name) _____ (Print Agency)

Signature of Person Receiving Instrument: _____

Date Receiving Instrument: ____/____/____ From: _____

Or Shipped Via:
USPS _____ UPS _____ FedEx _____ Other _____

Date Sending Instrument: ____/____/____ Air Bill No. _____

Complete Air Bill (Identify Instrument/Serial Number) and attach a copy to this worksheet.

Transported to the Crime Laboratory:

Agency Returning Instrument: CMI, Inc.

Name/Agency of Person Returning Instrument:

(Print Name) _____ (Print Agency)

Signature of Person Returning Instrument: _____

Date Returning Instrument: ____/____/____ Received By: _____

Or Received Via:
USPS _____ UPS _____ FedEx Other _____

Date Receiving Instrument: 21 July 2010 Air Bill No. No paperwork

Attach shipping paperwork (if any) to this worksheet.

Instrument Transfer Information Entered Into the Equipment Database:

Date of Action ^{21 DK} 22 July 2010 By: Deb Kashner

INTOX
PBT
TOXITEST

DATE 7-1-10
RECEIVED VIA FEDEX

CMI INC.

316 E. 9th St., Owensboro, KY 42303
270-685-6200 — Fax 270-685-6288
1-800-835-0690

SERVICE WORK ORDER

324333

B
I
L
L
T
O

ATTN:
CITY STATE ZIP
CUSTOMER NO. 585052
PHONE

S
H
I
P
T
O

Office of Attorney General
ATTN: Crime Lab Division
2641 E. Main St.
Marilyn Pearson
CITY STATE ZIP
Bismarck ND 58501
PHONE 701-328-6159

MODEL # 8000	S/N 80-004204	DATE RCVD.	P.O. #	REQ.	BILL CODE E W	SERVICE LOCATION CODE
-----------------	------------------	------------	--------	------	------------------	-----------------------

PROD. CODE	EXTRA PARTS RCVD. PC, BT, gas box, battery
------------	---

DESCRIPTION OF PROBLEM
"Invalid Res Returned"

QUOTE WARRANTY
 SERVICE DEPT. NEW

- PRELIMINARY TEST DISMANTLE, INSPECT, DIAGNOSTIC TEST
- POWER SUPPLY TEST, INSPECT REPLACE
- MAIN PWB's - REMOVE & CLEAN, REINSTALL, TEST
- PWB's DIAGNOSE, REPAIR REPLACE
- PWB's - SURFACE CLEAN IN PLACE
- FILTER WHEEL - REMOVE, CLEAN, REINSTALL, CHECK CELL INTERIOR
- CHOPPER WHEEL MOTOR TEST & INSPECT REPLACE
- CELL - REMOVE, CLEAN, REINSTALL, REPLACE LENSES/GASKETS REQ'D
- IR LIGHT SOURCE REMOVE, REALIGN REPLACE
- PRINTER ASSY. DISASSEMBLE CLEAN & LUBRICATE
- REMOVE & REPLACE BELT REALIGNMENT
- REPLACE TEST PRINTER FUNCTIONS
- BATTERIES - REPLACE CPU RAM
- REPLACE PARTS REASSEMBLE CALIBRATION FINAL TEST
- MISC.
- PBT PRELIMINARY TESTING - PRESSURE, BATTERY, ZERO SET
- DISASSEMBLE REPLACE PARTS REASSEMBLE
- TEST CALIBRATE FINAL TEST
- TOXITEST INCOMING LEAK TEST VISUAL/MECH. INSPECTION
- TEST THERMOMETER & THERMOSTAT DISASSEMBLE
- REPLACE PARTS REASSEMBLE OUTGOING & LEAK TEST
- FINAL TEST
- MISC.

PARTS USED			
QTY	ID NO.	DESCRIPTION	CODE
1	020982	Source Assy	

COMMENTS:
Replaced parts listed

S-01 = 03393

SHIPPING DATE/CHARGES
80-004204
P. 54 of 55
YH

HOURS: (4.0 hr)
TECH. K. R. B. DATE 7/13/10 CAL. TECH. LR DATE 7/14/10

FORM # CMI 8WO 1 REV (-)

SPECIAL SHIPPING INSTRUCTIONS
 COD AMOUNT \$

Certificate of Calibration

Intoxilyzer® 8000

*This is to certify that the calibration of **Intoxilyzer**® serial number 80-004204, manufactured by CMI Inc., a subsidiary of MPD, Inc. of Owensboro, Kentucky, was tested and found to be in compliance with the National Highway Traffic Safety Administration Standard for Devices to Measure Breath Alcohol (F.R., No. 179 48705-48710 Sept. 17, 1993) and with International Recommendation OIML R126 Edition 1998 (E). Calibration solutions are traceable to NIST standard material 1828.*

Date 7/14/10 Signed Kristi Robinson
Technician

CMI INC.

80-004204
P. 55 of 55
FM

316 East 9th Street
Owensboro, KY 42303
USA

Part No. 650519 CMI 7/9/09