

INTOXILYZER 8000 INSPECTION

Serial No.: 80-004199

Location: TOXL

Check When Done

- A. Is the warm-up time less than 20 minutes? ✓
- B. Is a three-pronged grounded outlet used? ✓
- C. Is the breath tube heated? ✓
- D. Is the diagnostic check complete? ✓
- E. Is the time, date, and year correct (re-set if necessary)? ✓
- F. Print test. (Attach test record.) ✓
- G. Low AC. Use <0.03 AC in ACA mode. (Attach test record.)
Sim SN DR5132 Lot No. Rep AC 0.021 ✓
- H. Linearity Test. Use ≥ 0.25 AC in ACA mode. (Attach test record.)
Sim SN DR5185 Lot No. Rep AC 0.297 ✓
- I. Interferent check. Use 0.05% acetone plus 0.10 AC ethanol in ABA mode.
(Attach test record.)
Sim SN DR3846 Lot No. Rep AC 0.10 AC + acetone ✓
- J. RFI check. Run CMS mode. Key radio on first room air.
(Attach test record.) ✓
- K. Calibration Check. Use Ethanol Breath Standard cylinder. Do three sets in ACA mode.
(Attach test record.)

Lot No. 11611080A2 Cyl No. 123 Exp. Date 06/01/2013
Test 1 0.080 Test 4 0.080 Test 7 0.081
Test 2 0.080 Test 5 0.081 Test 8 0.080
Test 3 0.080 Test 6 0.080 Test 9 0.081
Average 0.080

L. Tank pressure, Level 3 Func D Sub G P: 535 psi Reg 530 psi

M. Remarks/Maintenance record: annual inspection after an EEPROM change at CMI.

21 July 2011
Date

Deb Keshur
Field Inspector's Signature

7-26-11
Date

[Signature]
Reviewed by

80-004199
P. 132 of 140
YFM

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004199
Location = TOXL 8164.13.00 06/09
07/21/2011 16:50

***** Printer Test *****

abcdefghijklmnopqrstuvwxyz1234567890-=_|
ABCDEFGHIJKLMNPOQRSTUVWXYZ!@#\$\$%^&*()_+?

abcdefghijklmnopqrstuvwxyz1234567890-=_|
ABCDEFGHIJKLMNPOQRSTUVWXYZ!@#\$\$%^&*()_+?

Current Instrument Setup

Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? Yes
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 11611080A2
Standard Cyl #? 123
Standard Expiration? 06/01/2013
Oper No? 777777

Flow Cal. Date: 07/05/2011
Slope 672
Intercept -604945

IR Calibration Date: 07/05/2011

	3um	9um
0th Coef(*100):	-11921	-14558
1st Coef(*100):	259221	127392
2nd Coef(*100):	3824	1697
H2O adj(mg/l*10k):	486	640

***** Printer Test End *****

Deb Kashur

Operator Signature
DEB KASHUR

Remarks:

Print test

Form 106-I8000

80-004199
P. 133 of 140
AM

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004199
Location = TOXL 8164.13.00 06/09
07/21/2011 16:50

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	16:51
02 Std. Sol.	0.020	16:51
03 Room Air	0.000	16:52
04 Std. Sol.	0.020	16:53
05 Room Air	0.000	16:53
06 Std. Sol.	0.020	16:54
07 Room Air	0.000	16:54


08 Sim Temp = 34.0°C

Simul Ser No = 0

Std Sol No = 0

County = 08

Oper No. = 777777



Operator Signature
DEB KASHUR

Remarks:

Low AC 0.021 AC

Form 106-I8000

80-004199
P.134 of 140
JPM

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004199
Location = TOXL 8164.13.00 06/09
07/21/2011 16:55

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	16:55
02 Std. Sol.	0.294	16:56
03 Room Air	0.000	16:57
04 Std. Sol.	0.294	16:57
05 Room Air	0.000	16:58
06 Std. Sol.	0.293	16:59
07 Room Air	0.000	16:59

08 Sim Temp = 34.0°C

Simul Ser No = 0

Std Sol No = 0

County = 08

Oper No. = 777777

Deb Kasher
Operator Signature
DEB KASHUR

Remarks:

Linearity test 0.297AC

Form 106-I8000

*80-004199
P.135 of 140
JKM*

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004199
Location = TOXL 8164.13.00 06/09
07/21/2011 17:05

Test	AC	Time
01 Room Air	0.000	17:06
02 *Subject Test	INT*	17:06
03 Room Air	0.000	17:06

*Invalid Test
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 02/01/1992
Sub Sex = Female Weight = 150
Test = DUI Cit = NA
Dr. Lic. = ND/DIS921456
Lot No = 11611080A2
Cyl No = 123
Expiration Date = 06/01/2013
County = 08 Oper No. = 777777

Deb Kasher
Operator Signature
DEB KASHUR

Remarks:

Interferent check 0.10 AC + acetone

Form 106-I8000

80-004199
P. 136 of 140
KML

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004199
Location = TOXL 8164.13.00 06/09
07/21/2011 17:07

Test	AC	Time
01 Diagnostic	OK	17:07
02 Room Air	RFI*	17:08
03 Room Air	0.000	17:08

*Invalid Test
Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 02/01/1992
Sub Sex = Female Weight = 150
Test = DUI Cit = NA
Dr. Lic. = ND/DIS921456
Lot No = 11611080A2
Cyl No = 123
Expiration Date = 06/01/2013
County = 08 Oper No. = 777777

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.

Deb Kasher

Operator Signature
DEB KASHUR

Remarks:

RFI check

Form 106-I8000

80-004199
P.132 of 140
KM

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004199
Location = TOXL 8164.13.00 06/09
07/21/2011 17:08

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	17:09
02 Std. Gas	0.080	17:09
03 Room Air	0.000	17:10
04 Std. Gas	0.080	17:10
05 Room Air	0.000	17:11
06 Std. Gas	0.080	17:11
07 Room Air	0.000	17:12

Lot No = 11611080A2
Cyl No = 123
Exp Date = 06/01/2013
County = 08 Oper No. = 777777

Deb Kasher
Operator Signature
DEB KASHUR

Remarks:
Calibration check 0.080 AC

Form 106-I8000

*80-004199
P. 138 of 140
JRM*

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004199
Location = TOXL 8164.13.00 06/09
07/21/2011 17:12

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	17:13
02 Std. Gas	0.080	17:13
03 Room Air	0.000	17:13
04 Std. Gas	0.081	17:14
05 Room Air	0.000	17:14
06 Std. Gas	0.080	17:15
07 Room Air	0.000	17:15

Lot No = 11611080A2
Cyl No = 123
Exp Date = 06/01/2013
County = 08 Oper No. = 777777

Deb Keshur
Operator Signature
DEB KASHUR

Remarks:

Calibration check 0.080 AC

Form 106-I8000

*80-004199
p. 139 of 140
JRM*

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004199
Location = TOXL 8164.13.00 06/09
07/21/2011 17:15

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	17:16
02 Std. Gas	0.081	17:16
03 Room Air	0.000	17:17
04 Std. Gas	0.080	17:17
05 Room Air	0.000	17:18
06 Std. Gas	0.081	17:18
07 Room Air	0.000	17:19

Lot No = 11611080A2
Cyl No = 123
Exp Date = 06/01/2013
County = 08 Oper No. = 777777

Deb Kashur

Operator Signature
DEB KASHUR

Remarks:

Calibration check 0.080AC

Form 106-I8000

*80-004199
P. 140 of 140
JAN*