

INTOXILYZER 8000 INSPECTION

Serial No.: 80-003071

Location: TOXL

- |   | Check When Done |
|---|-----------------|
| A. Is the warm-up time less than 20 minutes?  | <u>✓</u>        |
| B. Is a three-pronged grounded outlet used?   | <u>✓</u>        |
| C. Is the breath tube heated?   | <u>✓</u>        |
| D. Is the diagnostic check complete?  | <u>✓</u>        |
| E. Is the time, date, and year correct (re-set if necessary)?   | <u>✓</u>        |
| F. Print test. (Attach test record.)  | <u>✓</u>        |
| G. Low AC. Use <0.03 AC in ACA mode. (Attach test record.)<br>Sim SN <u>DR5144</u> Lot No. <u>—</u> Rep AC <u>0.021</u>   | <u>✓</u>        |
| H. Linearity Test. Use ≥ 0.25 AC in ACA mode. (Attach test record.)<br>Sim SN <u>DR5114</u> Lot No. <u>—</u> Rep AC <u>0.298</u>                                | <u>✓</u>        |
| I. Interferent check. Use 0.05% acetone plus 0.10 AC ethanol in ABA mode.<br>(Attach test record.)<br>Sim SN <u>DR5186</u> Lot No. <u>—</u> Rep AC <u>0.100</u> | <u>✓</u>        |
| J. RFI check. Run CMS mode. Key radio on first room air.<br>(Attach test record.)   | <u>✓</u>        |
| K. Calibration Check. Use Ethanol Breath Standard cylinder. Do three sets in ACA mode.<br>(Attach test record.)   |                 |

Lot No. <u>709352</u>	Cyl No. <u>38</u>	Exp. Date <u>23 Dec 2011</u>
Test 1 <u>0.079</u>	Test 4 <u>0.079</u>	Test 7 <u>0.079</u>
Test 2 <u>0.078</u>	Test 5 <u>0.078</u>	Test 8 <u>0.079</u>
Test 3 <u>0.079</u>	Test 6 <u>0.078</u>	Test 9 <u>0.080</u>
Average <u>0.079</u>		

L. Tank pressure, Level 3 Func D Sub G P: 500 psi Reg 530 psi

M. Remarks/Maintenance record: Annual inspection

22 July 2010  
Date

Deb Kashur  
Field Inspector's Signature

29 July 2010  
Date

Manny Beson  
Reviewed by

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000            SN 80-003071  
Location = TOXL                    8164.13.00 06/09  
07/22/2010                            11:13

\*\*\*\*\* Printer Test \*\*\*\*\*

abcdefghijklmnopqrstuvwxy1234567890-=  
ABCDEFGHIJKLMNopqrstuvwxyz!@#%&^\*()\_+?

abcdefghijklmnopqrstuvwxy1234567890-=  
ABCDEFGHIJKLMNopqrstuvwxyz!@#%&^\*()\_+?

Current Instrument Setup

Data Entry Mode:            Enabled  
Start Test Sequence:        DABACABA  
Display Prelim Rslt?        Yes  
Display Third Digit?        Yes  
Inhib Printer(Y/N)?        No  
Display Volume?            No  
Disable On Memfull?        Yes  
# of Print Copies?          1  
Select Std (D/W/I)?        Dry  
Standard Value?            0.080  
Standard Lot #?            709352  
Standard Cyl #?            38  
Standard Expiration?        12/23/2011  
Oper No?                    777777

Flow Cal. Date:            07/09/2010  
Slope                        806  
Intercept                   -994737

IR Calibration Date:        07/09/2010

	3um	9um
0th Coef(*100):	-16015	-11994
1st Coef(*100):	278129	138369
2nd Coef(*100):	1927	1066
H2O adj(mg/l*10k):	293	137

\*\*\*\*\* Printer Test End \*\*\*\*\*

Deb Kashur

Operator Signature  
DEB KASHUR

Remarks:

Print test

Form 106-I8000

80-003071  
P. 9/1 of 9P  
JSM

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-003071  
Location = TOXL      8164.13.00 06/09  
07/22/2010      15:42

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	15:43
02 Std. Sol.	0.019	15:45
03 Room Air	0.000	15:45
04 Std. Sol.	0.020	15:46
05 Room Air	0.000	15:46
06 Std. Sol.	0.020	15:47
07 Room Air	0.000	15:47

08 Sim Temp = 34.0°C

Simul Ser No = 00000

Std Sol No = 000

County = 08

Oper No. = 777777

*Deb Kashur*

Operator Signature  
DEB KASHUR

Remarks:

*low Ac      0.021AC*

Form 106-I8000

*80-003071  
P. 92/98  
JK*

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-003071  
Location = TOXL      8164.13.00 06/09  
07/22/2010      15:49

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	15:49
02 Std. Sol.	0.301	15:50
03 Room Air	0.000	15:51
04 Std. Sol.	0.301	15:51
05 Room Air	0.000	15:52
06 Std. Sol.	0.300	15:53
07 Room Air	0.000	15:53

08 Sim Temp = 34.0°C

Simul Ser No = 00000

Std Sol No = 000

County = 08

Oper No. = 777777

*Deb Kashur*

Operator Signature  
DEB KASHUR

Remarks:

*Linearity test 0.298 AC*

Form 106-18000

80-003071  
P. 93 of 98  
FM

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000        SN 80-003071  
Location = TOXL                8164.13.00 06/09  
07/22/2010                        16:00

Test	AC	Time
01 Room Air	0.000	16:01
02 *Subject Test	INT*	16:01
03 Room Air	0.000	16:02

\*Invalid Test  
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 02/01/1992  
Sub Sex = Female                Weight = 150  
Test = DUI                        Cit = NA  
Dr. Lic. = ND/DIS921456  
Lot No = 709352  
Cyl No = 38  
Expiration Date = 12/23/2011  
County = 08                        Oper No. = 777777

*Deb Kashur*

Operator Signature  
DEB KASHUR

Remarks:

*Interferent check 0.100 AC + acetone*

Form 106-I8000

80-003071  
P. 94-198  
*[Signature]*

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000        SN 80-003071  
Location = TOXL                 8164.13.00 06/09  
07/22/2010                        16:10

Test	AC	Time
01 Diagnostic	OK	16:10
02 Room Air	RFI*	16:10
03 Room Air	0.000	16:11

\*Invalid Test  
Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 02/01/1992  
Sub Sex = Female                 Weight = 150  
Test = DUI                         Cit = NA  
Dr. Lic. = ND/DIS921456  
Lot No = 709352  
Cyl No = 38  
Expiration Date = 12/23/2011  
County = 08                        Oper No. = 777777

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.

Deb Kashur

Operator Signature  
DEB KASHUR

Remarks:

RFI check

Form 106-I8000

80-003071  
P. 95 of 98  
JK

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000        SN 80-003071  
Location = TOXL                 8164.13.00 06/09  
07/22/2010                        11:14

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:15
02 Std. Gas	0.079	11:15
03 Room Air	0.000	11:16
04 Std. Gas	0.078	11:16
05 Room Air	0.000	11:17
06 Std. Gas	0.079	11:17
07 Room Air	0.000	11:17

Lot No = 709352  
Cyl No = 38  
Exp Date = 12/23/2011  
County = 08                        Oper No. = 777777

Deb Kasher  
Operator Signature  
DEB KASHUR

Remarks:

*calibration check      0.080 AC*

Form 106-I8000

*80-003071  
P. 96 of 98  
fm*

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-003071  
Location = TOXL      8164.13.00 06/09  
07/22/2010      11:18

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:18
02 Std. Gas	0.079	11:19
03 Room Air	0.000	11:19
04 Std. Gas	0.078	11:20
05 Room Air	0.000	11:20
06 Std. Gas	0.078	11:20
07 Room Air	0.000	11:21

Lot No = 709352  
Cyl No = 38  
Exp Date = 12/23/2011  
County = 08      Oper No. = 777777

*Deb Kashur*

Operator Signature  
DEB KASHUR

Remarks:

*Calibration check 0.080 AC*

Form 106-I8000

80-003071  
P. 97 of 98  
JKM



Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000        SN 80-003071  
Location = TOXL                8164.13.00 06/09  
07/22/2010                        11:21

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:22
02 Std. Gas	0.079	11:22
03 Room Air	0.000	11:23
04 Std. Gas	0.079	11:23
05 Room Air	0.000	11:24
06 Std. Gas	0.080	11:24
07 Room Air	0.000	11:25

Lot No = 709352  
Cyl No = 38  
Exp Date = 12/23/2011  
County = 08                      Oper No. = 777777

*Deb Kasher*

Operator Signature  
DEB KASHUR

Remarks:

*Calibration check 0.080 AC*

Form 106-I8000

80-003071  
p. 98 of 98  
JK