



# ETHANOL BREATH STANDARD CYLINDER REPORT

OFFICE OF ATTORNEY GENERAL  
CRIME LABORATORY DIVISION  
Toxicology Section/Breath Alcohol Program  
SFN 59282 (5-2011)

Chemical Test Operator Name (Print) <i>JASON FOHLKAMP</i>		
Location <i>BURKE COUNTY SHERIFF'S</i>	Intoxilyzer® Serial Number <i>80-004940</i>	
Gas Lot Number <i>13314080A2</i>	Gas Cylinder Number <i>58</i>	Gas Expiration Date <i>06/05/16</i>

Check When Done:

- 1. Scan/Enter Gas Cylinder Information (Level 1, Function S).
- 2. Perform an ACA Test (Level 1, Function C).  
Write in the ACA Test Results (Reported to 3 Digits, Ex. 0.081).
  - A. *0.078* AC
  - B. *0.078* AC
  - C. *0.079* AC
- 3. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.
- 4. File Previous Intoxilyzer® Record (SFN50496, Form 120-G) at the Intoxilyzer® Location at the Agency.
- 5. Send the Following to the Crime Laboratory:
  - A. Ethanol Breath Standard Cylinder Report (SFN59282, Form 105-G).
  - B. ACA Test Record.

Chemical Test Operator Signature <i>[Signature]</i> #4171	Date <i>7/23/14</i>
Reviewed By (Crime Laboratory Use Only) <i>Deb Kashur</i>	Date <i>29 July 14</i>


Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004940  
Location = BURK      8164.13.00 06/09  
07/23/2014      13:36

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:37
02 Std. Gas	0.078	13:37
03 Room Air	0.000	13:38
04 Std. Gas	0.078	13:38
05 Room Air	0.000	13:38
06 Std. Gas	0.079	13:39
07 Room Air	0.000	13:39


Lot No = 13314080A2  
Cyl No = 58  
Exp Date = 06/05/2016  
County = 07      Oper No. = 130548

  
\_\_\_\_\_  
Operator Signature  
JASON POHLKAMP

Remarks:

Form 106-I8000

I certify that the attached is a true and correct copy of the Intoxilyzer test record.

Print Name and Badge # JASON POHLKAMP #4171  
Signature  #4171  
Date 7/23/14