



ETHANOL BREATH STANDARD CYLINDER REPORT
OFFICE OF ATTORNEY GENERAL
CRIME LABORATORY DIVISION
Toxicology Section/Breath Alcohol Program
SFN 59282 (1-2010)

Only Chemical Test Operators may change the Ethanol Breath Standard and complete this form.

Chemical Test Operator Name: <u>Mark W. McDonald</u>		
Location: <u>Valley City Police Dept.</u>		
Intoxilyzer® Serial Number: <u>40-004201</u>		
Gas Lot Number: <u>659358</u>	Gas Cylinder Number: <u>24</u>	Expiration Date: <u>9/3/2011</u>
<input type="checkbox"/> New Cylinder Installed	<input checked="" type="checkbox"/> Cylinder Not Changed	
Instrument Test Results (Report to 3 Digits; Example, 0.081%)		
Test 1: <u>0.077</u> %		
Test 2: <u>0.078</u> %		
Test 3: <u>0.077</u> %		
(ATTACH TEST RECORD)		

[Signature]
Chemical Test Operator Signature

7/29/2010
Date

Deb Kashner
Reviewed By

03 Aug 2010
Date

80-004201
P. 89-90
KM

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004201
Location = VALL 8164.13.00 06/09
07/24/2010 10:57

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:58
02 Std. Gas	0.077	10:59
03 Room Air	0.000	10:59
04 Std. Gas	0.078	11:00
05 Room Air	0.000	11:00
06 Std. Gas	0.077	11:00
07 Room Air	0.000	11:01

Lot No = 659358
Cyl No = 24
Exp Date = 09/03/2011
County = 02 Oper No. = 119502



Operator Signature
MARK W. MCDONALD

Remarks:

Form 106-18000

80-004201
P. 90 of 90
FM