



ETHANOL GAS STANDARD CYLINDER REPORT
 OFFICE OF ATTORNEY GENERAL
 CRIME LABORATORY DIVISION
 Toxicology Section/Breath Alcohol Program
 SFN 59282 (10-2013)

Chemical Test Operator Name (Print) <i>Andrew Eisenschenk</i>		
Location Valley City Police Department		Intoxilyzer® Serial Number <i>80-004191</i>
Gas Lot Number 16512080A2	Gas Cylinder Number 047	Gas Expiration Date 8/1/2014

Check When Done:

- 1. Scan/Enter Gas Cylinder Information (Level 1, Function S).
- 2. Perform an ACA Test (Level 1, Function C).
 Write in the ACA Test Results (Reported to 3 Digits, Ex. 0.081).
 A. *0.077* AC
 B. *0.078* AC
 C. *0.078* AC
- 3. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.
- 4. File Previous Intoxilyzer® Record (SFN50496, Form 120-G) at the Intoxilyzer® Location at the Agency.
- 5. Send the Following to the Crime Laboratory:
 - A. Ethanol Gas Standard Cylinder Report (SFN59282, Form 105-G).
 - B. ACA Test Record.

Chemical Test Operator Signature <i>[Signature]</i>	Date <i>7-24-14</i>
Reviewed By (Crime Laboratory Use Only) <i>Deb Kashner</i>	Date <i>24 July 14</i>

Moved across hall due to construction

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004191
Location = VALL 8164.13.00 06/09
07/24/2014 16:04

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	16:05
02 Std. Gas	0.077	16:05
03 Room Air	0.000	16:06
04 Std. Gas	0.078	16:06
05 Room Air	0.000	16:07
06 Std. Gas	0.078	16:07
07 Room Air	0.000	16:08

Lot No = 16512080A2

Cyl No = 47

Exp Date = 08/01/2014

County = 02

Oper No. = 130409



Operator Signature
ANDREW EISENSCHENK

Remarks:

Form 106-I8000