



ETHANOL GAS STANDARD CYLINDER REPORT

OFFICE OF ATTORNEY GENERAL

CRIME LABORATORY DIVISION

Toxicology Section/Breath Alcohol Program

SFN 59282 (10-2013)

Chemical Test Operator Name (Print) Daniel Garvens		
Location Pembina County Sheriff's Office		Intoxilyzer® Serial Number 80-004204
Gas Lot Number 13314080A3	Gas Cylinder Number 33	Gas Expiration Date 06/05/2016

Check When Done:

1. Scan/Enter Gas Cylinder Information (Level 1, Function S).

2. Perform an ACA Test (Level 1, Function C).

Write in the ACA Test Results (Reported to 3 Digits, Ex. 0.081).

A. 0. 0 8 0 AC

B. 0. 0 8 0 AC

C. 0. 0 8 0 AC

3. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.

4. File Previous Intoxilyzer® Record (SFN50496, Form 120-G) at the Intoxilyzer® Location at the Agency.

5. Send the Following to the Crime Laboratory:

A. Ethanol Gas Standard Cylinder Report (SFN59282, Form 105-G).

B. ACA Test Record.

Chemical Test Operator Signature 	Date 7-24-14
Reviewed By (Crime Laboratory Use Only) Deb Kashur	Date 28 July 14

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004204
Location = CAVA 8164.13.00 06/09
07/24/2014 13:54

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:54
02 Std. Gas	0.080	13:55
03 Room Air	0.000	13:55
04 Std. Gas	0.080	13:56
05 Room Air	0.000	13:56
06 Std. Gas	0.080	13:56
07 Room Air	0.000	13:57

Lot No = 13314080A3

Cyl No = 33

Exp Date = 06/05/2016

County = 34

Oper No. = 060849



Operator Signature
DANIEL GARVENS

Remarks:

New Gas Cylinder Install