



ETHANOL BREATH STANDARD CYLINDER REPORT

OFFICE OF ATTORNEY GENERAL
CRIME LABORATORY DIVISION
Toxicology Section/Breath Alcohol Program
SFN 59282 (1-2010)

Only Chemical Test Operators may change the Ethanol Breath Standard and complete this form.

Chemical Test Operator Name: <u>Chad L. Harmon</u>		
Location: <u>FYPD</u>		
Intoxilyzer® Serial Number: <u>80-004954</u>		
Gas Lot Number: <u>709352</u>	Gas Cylinder Number: <u>26</u>	Expiration Date: <u>12-23-11</u>
<input type="checkbox"/> New Cylinder Installed <input checked="" type="checkbox"/> Cylinder Not Changed		
Instrument Test Results (Report to 3 Digits; Example, 0.081%)		
Test 1: <u>0.078</u> %		
Test 2: <u>0.078</u> %		
Test 3: <u>0.078</u> %		
(ATTACH TEST RECORD)		

Chad L. Harmon
Chemical Test Operator Signature

07-25-10
Date

Debs Kashur
Reviewed By

29 July 2010
Date

80-004954
P. 57 of 58
KH

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004954
Location = FYPD 8164.13.00 06/09
07/25/2010 23:47

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	23:48
02 Std. Gas	0.078	23:48
03 Room Air	0.000	23:49
04 Std. Gas	0.078	23:49
05 Room Air	0.000	23:50
06 Std. Gas	0.078	23:50
07 Room Air	0.000	23:50

Lot No = 709352
Cyl No = 26
Exp Date = 12/23/2011
County = 61 Oper No. = 010606



Operator Signature
CHAD L. HARMON

Remarks:

Form 106-I8000

80-004954
P. 58 of 58
JHM