

CHAIN OF CUSTODY
BREATH ALCOHOL INSTRUMENT

Instrument: 8000 Serial Number: 80-005944

Agency being assigned instrument: Bismarck PD

Date 26 July 16

Transferred from:

Name (print) Debs Shanaver

Signature Dbs Shanaver

Agency Crime

Transferred to:

Name (print) Bismarck PD - Pat Benz

Signature _____

Agency _____

Reason for transfer/remarks:

Troubleshooting - both 5944 & their prior intox 5947 are reading Interferent Detected. The problem appears to be the air handling system.

28 July 16 - the air handling system has been flushed out & the ^{DS} 1st install is good.

Or Shipped Via:

USPS _____ UPS _____ FedEx _____ Other _____

Date Sending Instrument: _____ / _____ / _____ Air Bill No. _____