

CHAIN OF CUSTODY
BREATH ALCOHOL INSTRUMENT

Instrument: 8000 Serial Number: 80-005947

Agency being assigned instrument: Crime Lab

Date 26 July 16

Transferred from: (Pat Renz, BSPD)

Transferred to:

Name (print) Deb Shanaver

Name (print) Deb Shanaver

Signature D.b Shanaver

Signature D.b Shanaver

Agency Crime Lab

Agency Crime Lab

Reason for transfer/remarks:

Troubleshooting - this intox suddenly was giving Interferent
Detects & reading low ACA's. We thought it was due to a diabetic
subject but it turned out the problem was the newly installed
air handling system. There is nothing wrong w/ intox 5947.

Or Shipped Via:

USPS _____ UPS _____ FedEx _____ Other _____

Date Sending Instrument: _____ / _____ / _____ Air Bill No. _____