



ETHANOL BREATH STANDARD CYLINDER REPORT

OFFICE OF ATTORNEY GENERAL
CRIME LABORATORY DIVISION
Toxicology Section/Breath Alcohol Program
SFN 59282 (1-2010)

Only Chemical Test Operators may change the Ethanol Breath Standard and complete this form.

Chemical Test Operator Name: SHANNON P. DUISEN		
Location: TIOGA P.D.		
Intoxilyzer® Serial Number: 80-004938		
Gas Lot Number: 709352	Gas Cylinder Number: 21	Expiration Date: 12/31/11
<input type="checkbox"/> New Cylinder Installed	<input checked="" type="checkbox"/> Cylinder Not Changed	12/23/2011 DK
Instrument Test Results (Report to 3 Digits; Example, 0.081%)		
Test 1: 0.077 %		
Test 2: 0.076 %		
Test 3: 0.077 %		
(ATTACH TEST RECORD)		

SPD
Chemical Test Operator Signature

7/27/10
Date

Deb Kashner
Reviewed By

29 July 2010
Date

80-004938
p. 51 of 52
RSM

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004938
Location = TIOG 8164.13.00 06/09
07/27/2010 11:28

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:29
02 Std. Gas	0.077	11:29
03 Room Air	0.000	11:30
04 Std. Gas	0.076	11:30
05 Room Air	0.000	11:31
06 Std. Gas	0.077	11:31
07 Room Air	0.000	11:31

Lot No = 709352
Cyl No = 21
Exp Date = 12/23/2011
County = 53 Oper No. = 020632



Operator Signature
SHANNON P. DUISEN

Remarks: ACA

Form 106-I8000

80-004938
P 520/52
YAM