



# ETHANOL BREATH STANDARD CYLINDER REPORT

OFFICE OF ATTORNEY GENERAL  
CRIME LABORATORY DIVISION  
Toxicology Section/Breath Alcohol Program  
SFN 59282 (5-2011)

|  |  |                                       |
|--|--|---------------------------------------|
| Chemical Test Operator Name (Print)<br><i>Andrew Eisenschenk</i> |  |                                       |
| Location<br><i>Valley City Police Dept.</i>                      | Intoxilyzer® Serial Number<br><i>80-004191</i> |                                       |
| Gas Lot Number<br><i>33913080A4</i>                              | Gas Cylinder Number<br><i>19</i>               | Gas Expiration Date<br><i>11/2016</i> |

Check When Done:

- 1. Scan/Enter Gas Cylinder Information (Level 1, Function S).
- 2. Perform an ACA Test (Level 1, Function C).  
Write in the ACA Test Results (Reported to 3 Digits, Ex. 0.081).
  - A. *0.079* AC
  - B. *0.078* AC
  - C. *0.079* AC
- 3. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.
- 4. File Previous Intoxilyzer® Record (SFN50496, Form 120-G) at the Intoxilyzer® Location at the Agency.
- 5. Send the Following to the Crime Laboratory:
  - A. Ethanol Breath Standard Cylinder Report (SFN59282, Form 105-G).
  - B. ACA Test Record.

|   |                           |
|---|---------------------------|
| Chemical Test Operator Signature<br>                          | Date<br><i>7-27-14</i>    |
| Reviewed By (Crime Laboratory Use Only)<br><i>Deb Kashner</i> | Date<br><i>30 July 14</i> |

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004191  
Location = VALL      8164.13.00 06/09  
07/27/2014      14:33

DRY CAL CHECK

| Test        | AC    | Time  |
|-------------|-------|-------|
| 01 Room Air | 0.000 | 14:34 |
| 02 Std. Gas | 0.079 | 14:34 |
| 03 Room Air | 0.000 | 14:35 |
| 04 Std. Gas | 0.078 | 14:35 |
| 05 Room Air | 0.000 | 14:36 |
| 06 Std. Gas | 0.079 | 14:36 |
| 07 Room Air | 0.000 | 14:37 |

Lot No = 33913080A4

Cyl No = 19

Exp Date = 01/01/2016

County = 02

Oper No. = 130409



Operator Signature  
ANDREW EISENSCHENK

Remarks:

ALA CYLINDER REPLACEMENT

Form 106-I8000