



**ETHANOL BREATH STANDARD CYLINDER REPORT**  
 OFFICE OF ATTORNEY GENERAL  
 CRIME LABORATORY DIVISION  
 Toxicology Section/Breath Alcohol Program  
 SFN 59282 (1-2010)

Only Chemical Test Operators may change the Ethanol Breath Standard and complete this form.

Chemical Test Operator Name: <u>David M Wallace</u>		
Location: <u>Dickinson 281</u>		
Intoxilyzer® Serial Number: <u>80-004947</u>		
Gas Lot Number: <u>659358</u>	Gas Cylinder Number: <u>23</u>	Expiration Date: <u>9-3-11</u>
<input type="checkbox"/> New Cylinder Installed <input checked="" type="checkbox"/> Cylinder Not Changed		
Instrument Test Results (Report to 3 Digits; Example, 0.081%) Test 1: <u>0.079</u> % Test 2: <u>0.029</u> % Test 3: <u>0.078</u> %		
(ATTACH TEST RECORD)		

David M Wallace  
 Chemical Test Operator Signature

7-30-10  
 Date

Deb Kashner  
 Reviewed By

12 Aug 2010  
 Date

80-004947  
 Pg 47 of 48  
 BS


Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004947  
Location = DKPD      8164.13.00 06/09  
07/30/2010      11:12

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:13
02 Std. Gas	0.079	11:13
03 Room Air	0.000	11:14
04 Std. Gas	0.079	11:14
05 Room Air	0.000	11:15
06 Std. Gas	0.078	11:15
07 Room Air	0.000	11:15

Lot No = 659358  
Cyl No = 23  
Exp Date = 09/03/2011  
County = 45      Oper No. = 068332

  
Operator Signature  
DAVID M. WALLACE

Remarks:

Form 106-I8000

80-004947  
Pg 48 of 48  
28