



ETHANOL GAS STANDARD CYLINDER REPORT

OFFICE OF ATTORNEY GENERAL
CRIME LABORATORY DIVISION
Toxicology Section/Breath Alcohol Program
SFN 59282 (10-2013)

Chemical Test Operator Name (Print) <i>Richard D. Johnson</i>		
Location <i>Washburn</i>	Intoxilyzer® Serial Number <i>80-003057</i>	
Gas Lot Number <i>13314080A3</i>	Gas Cylinder Number <i>30</i>	Gas Expiration Date <i>06/05/16</i>

Check When Done:

- 1. Scan/Enter Gas Cylinder Information (Level 1, Function S).
- 2. Perform an ACA Test (Level 1, Function C).
Write in the ACA Test Results (Reported to 3 Digits, Ex. 0.081).
 - A. *0.078* AC
 - B. *0.078* AC
 - C. *0.079* AC
- 3. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.
- 4. File Previous Intoxilyzer® Record (SFN50496, Form 120-G) at the Intoxilyzer® Location at the Agency.
- 5. Send the Following to the Crime Laboratory:
 - A. Ethanol Gas Standard Cylinder Report (SFN59282, Form 105-G).
 - B. ACA Test Record.

Chemical Test Operator Signature <i>Richard D. Johnson</i>	Date <i>07/31/14</i>
Reviewed By (Crime Laboratory Use Only) <i>Deb Kashner</i>	Date <i>01 Aug 14</i>

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-003057
Location = WASH 8164.13.00 06/09
07/31/2014 15:50

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	15:51
02 Std. Gas	0.078	15:51
03 Room Air	0.000	15:52
04 Std. Gas	0.078	15:52
05 Room Air	0.000	15:53
06 Std. Gas	0.079	15:53
07 Room Air	0.000	15:53


Lot No = 13314080A3

Cyl No = 30

Exp Date = 06/05/2016

County = 28

Oper No. = 118707



Operator Signature
RICHARD JOHNSON

Remarks:

Form 106-I8000