



ETHANOL GAS STANDARD CYLINDER REPORT

OFFICE OF ATTORNEY GENERAL

CRIME LABORATORY DIVISION

Toxicology Section/Breath Alcohol Program

SFN 59282 (10-2013)

Chemical Test Operator Name (Print) <i>Gabriel Johnston</i>		
Location <i>Cavalier County Langdon</i>	Intoxilyzer® Serial Number <i>80-002667</i>	
Gas Lot Number <i>13314080A3</i>	Gas Cylinder Number <i>27</i>	Gas Expiration Date <i>06-05-16</i>

Check When Done:

1. Scan/Enter Gas Cylinder Information (Level 1, Function S).

2. Perform an ACA Test (Level 1, Function C).

Write in the ACA Test Results (Reported to 3 Digits, Ex. 0.081).

A. *0.081* AC

B. *0.080* AC

C. *0.080* AC

3. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.

4. File Previous Intoxilyzer® Record (SFN50496, Form 120-G) at the Intoxilyzer® Location at the Agency.

5. Send the Following to the Crime Laboratory:

A. Ethanol Gas Standard Cylinder Report (SFN59282, Form 105-G).

B. ACA Test Record.

Chemical Test Operator Signature <i>[Signature]</i>	Date <i>8-1-2014</i>
Reviewed By (Crime Laboratory Use Only) <i>[Signature]</i>	Date <i>07 Aug. 14</i>

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-002667
Location = LANG 8164.13.00 06/09
08/01/2014 19:40

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	19:41
02 Std. Gas	0.081	19:42
03 Room Air	0.000	19:42
04 Std. Gas	0.080	19:42
05 Room Air	0.000	19:43
06 Std. Gas	0.080	19:43
07 Room Air	0.000	19:44

Lot No = 13314080A3
Cyl No = 27
Exp Date = 06/05/2016
County = 10 Oper No. = 021009



Operator Signature
GABRIEL JOHNSTON

Remarks: