



ETHANOL BREATH STANDARD CYLINDER REPORT
 OFFICE OF ATTORNEY GENERAL
 CRIME LABORATORY DIVISION
 Toxicology Section/Breath Alcohol Program
 SFN 59282 (1-2010)

Only Chemical Test Operators may change the Ethanol Breath Standard and complete this form.

Chemical Test Operator Name: <u>GERARD R. ZLOTKOWSKI</u>		
Location: <u>ASHL</u>		
Intoxilyzer® Serial Number: <u>80-004945</u>		
Gas Lot Number: <u>709352</u>	Gas Cylinder Number: <u>32</u>	Expiration Date: <u>12/23/11</u>
<input type="checkbox"/> New Cylinder Installed <input checked="" type="checkbox"/> Cylinder Not Changed		
Instrument Test Results (Report to 3 Digits; Example, 0.081%) Test 1: <u>0.078</u> % Test 2: <u>0.079</u> % Test 3: <u>0.078</u> %		
(ATTACH TEST RECORD)		

Gerard R. Zlotkowski
 Chemical Test Operator Signature

8/2/2010
 Date

Deb Kashur
 Reviewed By

03 NOV 2010
 Date


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004945
Location = ASHL 8164.13.00 06/09
08/02/2010 17:02

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	17:03
02 Std. Gas	0.078	17:03
03 Room Air	0.000	17:04
04 Std. Gas	0.079	17:04
05 Room Air	0.000	17:05
06 Std. Gas	0.078	17:05
07 Room Air	0.000	17:05

Lot No = 709352
Cyl No = 32
Exp Date = 12/23/2011
County = 26 Oper No. = 100631



Operator Signature
GERARD R. ZLOTKOWSKI

Remarks:

Form 106-I8000

80-004945
P. 50 of 50
JAM