



OFFICE OF ATTORNEY GENERAL  
 Crime Laboratory Division  
 2641 East Main Avenue  
 P.O. Box 937  
 Bismarck, North Dakota 58502-0937

TEL 701-328-6159

**ETHANOL BREATH STANDARD CYLINDER INSTALLATION FORM**

(Only Chemical Test Operators may change the Ethanol Breath Standard and complete this form.)

Chemical Test Operator Name: Darcy W Aberle

Location: Williston LEC

Intoxilyzer Serial No.: 80-004189

Lot Number: 659357

Cylinder Number: 45

Expiration Date: 09/03/11

Instrument Test Results (Report to 3 Digits; Example, 0.081%)

Test 1: 0.076 %

Test 2: 0.077 %

Test 3: 0.077 %

(ATTACH TEST RECORD)

Darcy W Aberle  
 Chemical Test Operator Signature

08-03-10  
 Date Cylinder Changed or Tested

Form 105-G  
 Toxicology Laboratory  
 Crime Laboratory Division

Deb Kashner

04 Aug 10

80-004189  
P. 87 of 88  
km


Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004189  
Location = WILL      8164.13.00 06/09  
08/03/2010      08:43

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	08:44
02 Std. Gas	0.076	08:44
03 Room Air	0.000	08:45
04 Std. Gas	0.077	08:45
05 Room Air	0.000	08:45
06 Std. Gas	0.077	08:46
07 Room Air	0.000	08:46

Lot No = 659357  
Cyl No = 45  
Exp Date = 09/03/2011  
County = 53      Oper No. = 010101

  
Operator Signature  
DARCY W. ABERLE

Remarks:

Form 106-18000

80-004189  
P. 88 of 88  
KM