



# ETHANOL BREATH STANDARD CYLINDER REPORT

OFFICE OF ATTORNEY GENERAL  
CRIME LABORATORY DIVISION  
Toxicology Section/Breath Alcohol Program  
SFN 59282 (5-2011)

Chemical Test Operator Name (Print) <i>RAY Weber</i>		
Location <i>TRAIL CO. S.O.</i>	Intoxilyzer® Serial Number <i>80-004191</i>	
Gas Lot Number <i>23411080A1</i>	Gas Cylinder Number <i>37</i>	Gas Expiration Date <i>10/01/13</i>

Check When Done:

- 1. Scan/Enter Gas Cylinder Information (Level 1, Function S).
- 2. Perform an ACA Test (Level 1, Function C).  
Write in the ACA Test Results (Reported to 3 Digits, Ex. 0.081).  
A. 0 078 AC  
B. 0 078 AC  
C. 0 079 AC
- 3. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.
- 4. File Previous Intoxilyzer® Record (SFN50496, Form 120-G) at the Intoxilyzer® Location at the Agency.
- 5. Send the Following to the Crime Laboratory:
  - A. Ethanol Breath Standard Cylinder Report (SFN59282, Form 105-G).
  - B. ACA Test Record.

Chemical Test Operator Signature <i>Ray Weber</i>	Date <i>08/05/2012</i>
Reviewed By (Crime Laboratory Use Only) <i>Deb Kashur</i>	Date <i>09 Aug. 2012</i>

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004191  
Location = HILL      8164.13.00 06/09  
08/05/2012      23:12

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	23:13
02 Std. Gas	0.078	23:14
03 Room Air	0.000	23:14
04 Std. Gas	0.078	23:15
05 Room Air	0.000	23:15
06 Std. Gas	0.079	23:16
07 Room Air	0.000	23:16

Lot No = 23411080A1  
Cyl No = 37  
Exp Date = 10/01/2013  
County = 49      Oper No. = 058405

  
\_\_\_\_\_  
Operator Signature  
RAY G. WEBER

Remarks: