



# ETHANOL BREATH STANDARD CYLINDER REPORT

OFFICE OF ATTORNEY GENERAL  
CRIME LABORATORY DIVISION  
Toxicology Section/Breath Alcohol Program  
SFN 59282 (1-2010)

Only Chemical Test Operators may change the Ethanol Breath Standard and complete this form.

|   |  |                                    |
|---|--|------------------------------------|
| Chemical Test Operator Name: <u>Lt. Tracy L. Meidinger #258</u> |  |                                    |
| Location: <u>UNDPD</u>  |  |                                    |
| Intoxilyzer® Serial Number: <u>80-004934</u>                    |  |                                    |
| Gas Lot Number: <u>659358</u>                                   | Gas Cylinder Number: <u>31</u>                           | Expiration Date: <u>09-30-2011</u> |
| <input type="checkbox"/> New Cylinder Installed                 | <input checked="" type="checkbox"/> Cylinder Not Changed |                                    |
| Instrument Test Results (Report to 3 Digits; Example, 0.081%)   |  |                                    |
| Test 1: 0. <u>079</u> %   |  |                                    |
| Test 2: 0. <u>079</u> %   |  |                                    |
| Test 3: 0. <u>079</u> %   |  |                                    |
| (ATTACH TEST RECORD)  |  |                                    |

Lt. Tracy L. Meidinger  
Chemical Test Operator Signature

08-07-10  
Date

Deb Kashner  
Reviewed By

12 Aug 2010  
Date

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004934  
Location = UNDP      8164.13.00 06/09  
08/07/2010      18:56

DRY CAL CHECK

| Test        | AC    | Time  |
|-------------|-------|-------|
| 01 Room Air | 0.000 | 18:57 |
| 02 Std. Gas | 0.079 | 18:57 |
| 03 Room Air | 0.000 | 18:58 |
| 04 Std. Gas | 0.079 | 18:58 |
| 05 Room Air | 0.000 | 18:58 |
| 06 Std. Gas | 0.079 | 18:59 |
| 07 Room Air | 0.000 | 18:59 |

Lot No = 659358  
Cyl No = 31  
Exp Date = 09/03/2011  
County = 18      Oper No. = 119409

*Tracy L. Meidinger #258*  
Operator Signature  
TRACY L. MEIDINGER

Remarks:

Form 106-I8000

ACA 8-7-10

80-004934  
P. 53 of 53  
JFM