



ETHANOL BREATH STANDARD CYLINDER REPORT
OFFICE OF ATTORNEY GENERAL
CRIME LABORATORY DIVISION
Toxicology Section/Breath Alcohol Program
SFN 59282 (1-2010)

Only Chemical Test Operators may change the Ethanol Breath Standard and complete this form.

Chemical Test Operator Name: <u>Joseph Knowski</u>		
Location: <u>Devils Lake LEC</u>		
Intoxilyzer® Serial Number: <u>80-004193</u>		
Gas Lot Number: <u>659358</u>	Gas Cylinder Number: <u>20</u>	Expiration Date: <u>9-30-11</u>
<input type="checkbox"/> New Cylinder Installed	<input checked="" type="checkbox"/> Cylinder Not Changed	
Instrument Test Results (Report to 3 Digits; Example, 0.081%)		
Test 1: 0. <u>079</u> %		
Test 2: 0. <u>079</u> %		
Test 3: 0. <u>078</u> %		
(ATTACH TEST RECORD)		

Joseph Knowski #287
Chemical Test Operator Signature
Deb Kashur
Reviewed By

8-8-2010.
Date
30 NOV 2010
Date


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004193
Location = DEVL 8164.13.00 06/09
08/08/2010 17:59

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	18:00
02 Std. Gas	0.079	18:00
03 Room Air	0.000	18:01
04 Std. Gas	0.079	18:01
05 Room Air	0.000	18:02
06 Std. Gas	0.078	18:02
07 Room Air	0.000	18:02

Lot No = 659358
Cyl No = 20
Exp Date = 09/03/2011
County = 36 Oper No. = 109208


Operator Signature
JOSEPH J. KNOWSKI

Remarks:

Form 106-I8000

80-004193
P. 86 of 86
JFM