



ETHANOL GAS STANDARD CYLINDER REPORT

OFFICE OF ATTORNEY GENERAL
CRIME LABORATORY DIVISION
Toxicology Section/Breath Alcohol Program
SFN 59282 (10-2013)

Chemical Test Operator Name (Print) <i>Jeremy Grohs</i>		
Location <i>Burke Co. Sheriff's Office</i>	Intoxilyzer® Serial Number <i>80-004942</i>	
Gas Lot Number <i>19615080AS</i>	Gas Cylinder Number <i>18</i>	Gas Expiration Date <i>9-5-2017</i>

Check When Done:

- 1. Scan/Enter Gas Cylinder Information (Level 1, Function S).
- 2. Perform an ACA Test (Level 1, Function C).
Write in the ACA Test Results (Reported to 3 Digits, Ex. 0.081).
 - A. 0.081 AC
 - B. 0.080 AC
 - C. 0.080 AC
- 3. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.
- 4. File Previous Intoxilyzer® Record (SFN50496, Form 120-G) at the Intoxilyzer® Location at the Agency.
- 5. Send the Following to the Crime Laboratory:
 - A. Ethanol Gas Standard Cylinder Report (SFN59282, Form 105-G).
 - B. ACA Test Record.

Chemical Test Operator Signature <i>Jeremy Grohs</i>	Date <i>8/8/16</i> ✓
Reviewed By (Crime Laboratory Use Only) <i>Deb Shanaver</i>	Date <i>10 Aug 16</i>

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004942
Location = BURK 8164.13.00 06/09
08/08/2016 11:36


DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:38
02 Std. Gas	0.081	11:38
03 Room Air	0.000	11:39
04 Std. Gas	0.080	11:39
05 Room Air	0.000	11:40
06 Std. Gas	0.080	11:40
07 Room Air	0.000	11:41

Lot No = 19615080A5
Cyl No = 18
Exp Date = 09/05/2017
County = 07 Oper No. = 130230

I certify that the attached is a true and correct copy of the Intoxilyzer test record.

Print Name and Badge # Jeremy Grohs #4170
Signature 
Date 8/8/16


Operator Signature
JEREMY GROHS

Remarks:

Good Test