



ETHANOL BREATH STANDARD CYLINDER REPORT
 OFFICE OF ATTORNEY GENERAL
 CRIME LABORATORY DIVISION
 Toxicology Section/Breath Alcohol Program
 SFN 59282 (1-2010)

Only Chemical Test Operators may change the Ethanol Breath Standard and complete this form.

Chemical Test Operator Name: <u>David J. Zeis</u>		
Location: <u>Cramer County Sheriff's Office - Langdon, ND</u>		
Intoxilyzer® Serial Number: <u>80-004188</u>		
Gas Lot Number: <u>659358</u>	Gas Cylinder Number: <u>53</u>	Expiration Date: <u>09/03/11</u>
<input type="checkbox"/> New Cylinder Installed <input checked="" type="checkbox"/> Cylinder Not Changed		
Instrument Test Results (Report to 3 Digits; Example, 0.081%)		
Test 1: 0. <u>075</u> %		
Test 2: 0. <u>076</u> %		
Test 3: 0. <u>075</u> %		
(ATTACH TEST RECORD)		

David J. Zeis
 Chemical Test Operator Signature

08/09/10
 Date

Deb Kashner
 Reviewed By

13 Aug 2010
 Date

80-004188
 P. 9/10/92
 KM


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004188
Location = LANG 8164.13.00 06/09
08/09/2010 16:14

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	16:15
02 Std. Gas	0.075	16:15
03 Room Air	0.000	16:15
04 Std. Gas	0.076	16:16
05 Room Air	0.000	16:16
06 Std. Gas	0.075	16:17
07 Room Air	0.000	16:17

Lot No = 659358
Cyl No = 53
Exp Date = 09/03/2011
County = 10 Oper No. = 067518



Operator Signature
DAVID J. ZEIS

Remarks:

Form 106-I8000

80-004188
p. 92 of 92
FM