



**ETHANOL BREATH STANDARD CYLINDER REPORT**  
 OFFICE OF ATTORNEY GENERAL  
 CRIME LABORATORY DIVISION  
 Toxicology Section/Breath Alcohol Program  
 SFN 59282 (1-2010)

Only Chemical Test Operators may change the Ethanol Breath Standard and complete this form.

Chemical Test Operator Name: <u>Ray Weber</u>		
Location: <u>TRAIL CO. S.O.</u>		
Intoxilyzer® Serial Number: <u>80-004200</u>		
Gas Lot Number: <u>659358</u>	Gas Cylinder Number: <u>#45</u>	Expiration Date: <u>09-03-11</u>
<input type="checkbox"/> New Cylinder Installed	<input type="checkbox"/> Cylinder Not Changed	
Instrument Test Results (Report to 3 Digits; Example, 0.081%)		
Test 1: 0. <u>078</u> %		
Test 2: 0. <u>078</u> %		
Test 3: 0. <u>078</u> %		
(ATTACH TEST RECORD)		

Ray Weber  
 Chemical Test Operator Signature

8-10-2010  
 Date

Deb Kashner  
 Reviewed By

23 Aug 2010  
 Date

80-004200  
 P. 89 of 90  
 KM

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004200  
Location = HILL      8164.13.00 06/09  
08/10/2010      23:23

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	23:24
02 Std. Gas	0.078	23:25
03 Room Air	0.000	23:25
04 Std. Gas	0.078	23:25
05 Room Air	0.000	23:26
06 Std. Gas	0.078	23:26
07 Room Air	0.000	23:27

Lot No = 659358  
Cyl No = 45  
Exp Date = 09/03/2011  
County = 49      Oper No. = 058405

*Ray Weber*

Operator Signature  
RAY G. WEBER

Remarks:

Form 106-I8000

*ACA TEST*

80-004200  
p. 90 of 90  
van