




ETHANOL BREATH STANDARD CYLINDER REPORT

OFFICE OF ATTORNEY GENERAL
CRIME LABORATORY DIVISION
Toxicology Section/Breath Alcohol Program
SFN 59282 (1-2010)

Only Chemical Test Operators may change the Ethanol Breath Standard and complete this form.

Chemical Test Operator Name: <u>Adam Krabbenhoft</u>		
Location: <u>LaMoure County Courthouse</u>		
Intoxilyzer® Serial Number: <u>80-004942</u>		
Gas Lot Number: <u>709352</u>	Gas Cylinder Number: <u>14</u>	Expiration Date: <u>12-23-11</u>
<input type="checkbox"/> New Cylinder Installed	<input checked="" type="checkbox"/> Cylinder Not Changed	
Instrument Test Results (Report to 3 Digits; Example, 0.081%)		
Test 1: 0. <u>078</u> %		
Test 2: 0. <u>079</u> %		
Test 3: 0. <u>077</u> %		
(ATTACH TEST RECORD)		


Chemical Test Operator Signature

8-10-10
Date

Deb Kasher
Reviewed By

13 Aug 2010
Date

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004942
Location = LAMO 8164.13.00 06/09
08/10/2010 10:27

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:28
02 Std. Gas	0.078	10:29
03 Room Air	0.000	10:29
04 Std. Gas	0.079	10:29
05 Room Air	0.000	10:30
06 Std. Gas	0.077	10:30
07 Room Air	0.000	10:31


Lot No = 709352

Cyl No = 14

Exp Date = 12/23/2011

County = 23

Oper No. = 020828



Operator Signature
ADAM R. KRABBENHOFT

Remarks:

ACA Test

Form 106-18000

80-004942
p. 48 of 48
JRM