

**INTOXILYZER 8000  
REPAIR AND MAINTENANCE**

Serial No.: 80-003066

Location: TOXL

Check When Done

1. Is a three-pronged grounded outlet used? ✓
2. Is the warm-up time less than 20 minutes? ✓
3. Is the breath tube heated? ✓
4. Is the diagnostic check complete? ✓
5. Is the time, date, and year correct (re-set if necessary)? ✓
6. Print test. (Attach test record.) ✓
7. Calibration Check. Use Ethanol Breath Standard (EBS). Do three sets in ACA mode. (Attach test records.)

Lot No. 709 352 Cylinder No. 38 Expir. Date 23 Dec 2011

Test 1 0.079 Test 4 0.079 Test 7 0.079

Test 2 0.079 Test 5 0.079 Test 8 0.079

Test 3 0.078 Test 6 0.079 Test 9 0.080

Average 0.079

8. Low AC. Use < 0.03 AC in ACA mode. (Attach test record.) ✓
9. Linearity Test. Use  $\geq 0.25$  AC in ACA mode. (Attach test record.) ✓
10. Interferent check. Use 0.05% acetone plus 0.10 AC ethanol in ABA mode. (Attach test record.) ✓
11. RFI check. Run CMS mode. Key radio on first room air. (Attach test record.) ✓

80-003066

P. 110 of 119  
FM

Electrical Repairs:

1. Was the calibration adjusted/performed? (Attach test record.) yes, at CM1
2. Was the light source changed? NO
3. Was the sample chamber cleaned? NO

Other Maintenance/Repairs:

1. Was the printer cleaned/repaired? NO
2. Was the solenoid cleaned? NO
3. Was new tubing installed? NO
4. Were additional parts replaced/changed/repaired? YES
5. Remarks/Maintenance record: power supply was replaced at CM1 and recalibrated.

11 Aug 2010  
Date

12 Aug 2010  
Date

Deb Kashner  
Field Inspector's Signature

Mary Pedron  
Reviewed by



Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-003066  
Location = TOXL      8164.13.00 06/09  
08/11/2010      12:56

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	12:57
02 Std. Sol.	0.019	12:58
03 Room Air	0.000	12:58
04 Std. Sol.	0.019	12:59
05 Room Air	0.000	12:59
06 Std. Sol.	0.019	13:00
07 Room Air	0.000	13:01

08 Sim Temp = 34.0°C

Simul Ser No = 00000

Std Sol No = 000

County = 08

Oper No. = 777777

*Deb Kashur*

Operator Signature  
DEB KASHUR

Remarks:

*Low AC      0.021 AC*

Form 106-I8000

80-003066  
p. 113 of 119  
JK

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-003066  
Location = TOXL      8164.13.00 06/09  
08/11/2010      13:15

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:16
02 Std. Sol.	0.292	13:16
03 Room Air	0.000	13:17
04 Std. Sol.	0.293	13:18
05 Room Air	0.000	13:19
06 Std. Sol.	0.293	13:19
07 Room Air	0.000	13:20

08 Sim Temp = 34.0°C

Simul Ser No = 00000

Std Sol No = 000

County = 08

Oper No. = 777777

*Deb Kashur*

Operator Signature  
DEB KASHUR

Remarks:

*Linearity test 0.298AC*

Form 106-18000

80-003066  
p. 114 of 119  
AK

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-003066  
Location = TOXL      8164.13.00 06/09  
08/11/2010      14:21

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:21
02 Std. Gas	0.079	14:22
03 Room Air	0.000	14:22
04 Std. Gas	0.079	14:23
05 Room Air	0.000	14:23
06 Std. Gas	0.078	14:23
07 Room Air	0.000	14:24

Lot No = 709352  
Cyl No = 38  
Exp Date = 12/23/2011  
County = 08      Oper No. = 777777

*Deb Kasher*

Operator Signature  
DEB KASHUR

Remarks:

*calibration check 0.080 AC*

Form 106-I8000

80-003066  
P. 115 of 119  
JK

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-003066  
Location = TOXL      8164.13.00 06/09  
08/11/2010      14:24

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:25
02 Std. Gas	0.079	14:25
03 Room Air	0.000	14:26
04 Std. Gas	0.079	14:26
05 Room Air	0.000	14:27
06 Std. Gas	0.079	14:27
07 Room Air	0.000	14:28

Lot No = 709352  
Cyl No = 38  
Exp Date = 12/23/2011  
County = 08      Oper No. = 777777

*Deb Kasher*

Operator Signature  
DEB KASHUR

Remarks:

*Calibration check 0.080 AC*

Form 106-I8000

80-003066  
p. 116 of 119  
JKH

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-003066  
Location = TOXL      8164.13.00 06/09  
08/11/2010      14:28

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:29
02 Std. Gas	0.079	14:29
03 Room Air	0.000	14:29
04 Std. Gas	0.079	14:30
05 Room Air	0.000	14:30
06 Std. Gas	0.080	14:31
07 Room Air	0.000	14:31

Lot No = 709352  
Cyl No = 38  
Exp Date = 12/23/2011  
County = 08      Oper No. = 777777

*Deb Kasher*

Operator Signature  
DEB KASHUR

Remarks:

*Calibration check 0.080AC*

Form 106-I8000

80-003066  
P. 117 of 119  
JKM



Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000        SN 80-003066  
Location = TOXL                8164.13.00 06/09  
08/11/2010                      14:35

Test	AC	Time
01 Room Air	0.000	14:36
02 *Subject Test	INT*	14:36
03 Room Air	0.000	14:37

\*Invalid Test  
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 02/01/1992  
Sub Sex = Female                Weight = 150  
Test = DUI                        Cit = NA  
Dr. Lic. = ND/DIS921456  
Lot No = 709352  
Cyl No = 38  
Expiration Date = 12/23/2011  
County = 08                      Oper No. = 777777

*Deb Kasher*

Operator Signature  
DEB KASHUR

Remarks:

*Interferent check 0.100Ac + acetone*

Form 106-I8000

80-003066  
P. 118 of 119  
FM

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000        SN 80-003066  
Location = TOXL                8164.13.00 06/09  
08/11/2010                        14:33

Test	AC	Time
01 Diagnostic	OK	14:34
02 Room Air	RFI*	14:34
03 Room Air	0.000	14:34

\*Invalid Test  
  Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 02/01/1992  
Sub Sex = Female                Weight = 150  
Test = DUI                        Cit = NA  
Dr. Lic. = ND/DIS921456  
Lot No = 709352  
Cyl No = 38  
Expiration Date = 12/23/2011  
County = 08                        Oper No. = 777777

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.

Deb Kasher

Operator Signature  
DEB KASHUR

Remarks:

*RFI check*

Form 106-I8000

80-003066  
p. 119 of 119  
KMU