



**ETHANOL BREATH STANDARD CYLINDER REPORT**  
 OFFICE OF ATTORNEY GENERAL  
 CRIME LABORATORY DIVISION  
 Toxicology Section/Breath Alcohol Program  
 SFN 59282 (1-2010)

Only Chemical Test Operators may change the Ethanol Breath Standard and complete this form.

Chemical Test Operator Name: <u>Troy Hischer</u>		
Location: <u>Cass Co. Jail</u>		
Intoxilyzer® Serial Number: <u>80-004946</u>		
Gas Lot Number: <u>608180</u>	Gas Cylinder Number: <u>76</u>	Expiration Date: <u>9-21-11</u>
<input checked="" type="checkbox"/> New Cylinder Installed	<input type="checkbox"/> Cylinder Not Changed	
Instrument Test Results (Report to 3 Digits; Example, 0.081%)		
Test 1: <u>0.076</u> %		
Test 2: <u>0.078</u> %		
Test 3: <u>0.076</u> %		
(ATTACH TEST RECORD)		

Troy Hischer  
 Chemical Test Operator Signature

8-11-10  
 Date

Deb Kashur  
 Reviewed By

23 Aug 2010  
 Date

80-004946  
 p. 55 of 56  
 yam


Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004946  
Location = CASS      8164.13.00 06/09  
08/11/2010      13:46

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:46
02 Std. Gas	0.076	13:47
03 Room Air	0.000	13:47
04 Std. Gas	0.078	13:48
05 Room Air	0.000	13:48
06 Std. Gas	0.076	13:49
07 Room Air	0.000	13:49

Lot No = 668180  
Cyl No = 76  
Exp Date = 09/21/2011  
County = 09      Oper No. = 049524

  
Operator Signature  
TROY A. HISCHER

Remarks:

Form 106-I8000

80-004946  
P. 56 of 56  
YAM