



# ETHANOL BREATH STANDARD CYLINDER REPORT

OFFICE OF ATTORNEY GENERAL  
CRIME LABORATORY DIVISION  
Toxicology Section/Breath Alcohol Program  
SFN 59282 (1-2010)

Only Chemical Test Operators may change the Ethanol Breath Standard and complete this form.

Chemical Test Operator Name: <u>Danielle M. Rosewarren</u>		
Location: <u>CARR</u>		
Intoxilyzer® Serial Number: <u>80-004944</u>		
Gas Lot Number: <u>709352</u>	Gas Cylinder Number: <u>12</u>	Expiration Date: <u>12/23/11</u>
<input type="checkbox"/> New Cylinder Installed <input checked="" type="checkbox"/> Cylinder Not Changed		
Instrument Test Results (Report to 3 Digits; Example, 0.081%)		
Test 1: 0. <u>079</u> %		
Test 2: 0. <u>079</u> %		
Test 3: 0. <u>080</u> %		
(ATTACH TEST RECORD)		

Danielle Rosewarren  
Chemical Test Operator Signature

08/12/10  
Date

Deb Kashner  
Reviewed By

13 Aug 2010  
Date

80-004944  
p. 51 of 52  
ym


Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000        SN 80-004944  
Location = CARR                8164.13.00 06/09  
08/12/2010                        15:16

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	15:16
02 Std. Gas	0.079	15:17
03 Room Air	0.000	15:17
04 Std. Gas	0.079	15:18
05 Room Air	0.000	15:18
06 Std. Gas	0.080	15:19
07 Room Air	0.000	15:19

Lot No = 709352  
Cyl No = 12  
Exp Date = 12/23/2011  
County = 16                      Oper No. = 100754



Operator Signature  
DANIELLE M. ROSEWAREN

Remarks:

Form 106-I8000

80-004944  
p. 52 of 52  
jm