

OFFICE OF ATTORNEY GENERAL
CRIME LABORATORY DIVISION
Toxicology Section/Breath Alcohol Program

Form: BrF-105-G

ETHANOL BREATH STANDARD CYLINDER REPORT

(Only Chemical Test Operators may change the Ethanol Breath Standard and complete this form.)

Chemical Test Operator Name: Steven R Fischer

Location: Rolette County Sheriff's Office, Rolla, ND

Intoxilyzer Serial No.: 80-004199

Lot Number: 659358

Cylinder Number: 11

Expiration Date: 9-3-11

Instrument Test Results (Report to 3 Digits; Example, 0.081%)

Test 1: 0.080%

Test 2: 0.080%

Test 3: 0.080%

(ATTACH TEST RECORD)

[Signature]
Chemical Test Operator Signature

8-15-10
Date

Deb Kashner
Reviewed By

23 Aug 2010
Date

80-00 4199
p. 89 of 90
JRM


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004199
Location = ROLL 8164.13.00 06/09
08/15/2010 05:04

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	05:05
02 Std. Gas	0.080	05:05
03 Room Air	0.000	05:05
04 Std. Gas	0.080	05:06
05 Room Air	0.000	05:06
06 Std. Gas	0.080	05:07
07 Room Air	0.000	05:07

Lot No = 659358
Cyl No = 11
Exp Date = 09/03/2011
County = 40 Oper No. = 049904



Operator Signature
STEVEN R. FISCHER

Remarks:

Form 106-I8000

80-004199
P. 90 of 90
JAN