



**ETHANOL BREATH STANDARD CYLINDER REPORT**  
 OFFICE OF ATTORNEY GENERAL  
 CRIME LABORATORY DIVISION  
 Toxicology Section/Breath Alcohol Program  
 SFN 59282 (1-2010)

Only Chemical Test Operators may change the Ethanol Breath Standard and complete this form.

Chemical Test Operator Name:	Todd A. Leroi		
Location:	GRAF		
Intoxilyzer® Serial Number:	80-003072		
Gas Lot Number:	<del>16</del> 659358 <sup>DK</sup>	Gas Cylinder Number:	<del>659358-16</del> <sup>DK</sup> Expiration Date: 09/03/11
<input type="checkbox"/> New Cylinder Installed	<input checked="" type="checkbox"/> Cylinder Not Changed		
Instrument Test Results (Report to 3 Digits; Example, 0.081%)			
Test 1: 0.082 %			
Test 2: 0.081 %			
Test 3: 0.082 %			
(ATTACH TEST RECORD)			

T. A. Leroi  
 Chemical Test Operator Signature

08/16/2010  
 Date

Deb Keshur  
 Reviewed By

23 Aug 2010  
 Date

80-003072  
 P. 91/492  
 YAM


Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-003072  
Location = GRAF      8164.13.00 06/09  
08/16/2010      22:50

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	22:50
02 Std. Gas	0.082	22:51
03 Room Air	0.000	22:51
04 Std. Gas	0.081	22:52
05 Room Air	0.000	22:52
06 Std. Gas	0.082	22:52
07 Room Air	0.000	22:53

Lot No = 659358  
Cyl No = 16  
Exp Date = 09/03/2011  
County = 50      Oper No. = 019709



Operator Signature  
TODD A. LEROL

Remarks:

Form 106-I8000

80-003072  
P. 92 of 92  
JAM