



**ETHANOL BREATH STANDARD CYLINDER REPORT**  
 OFFICE OF ATTORNEY GENERAL  
 CRIME LABORATORY DIVISION  
 Toxicology Section/Breath Alcohol Program  
 SFN 59262 (1-2010)

Only Chemical Test Operators may change the Ethanol Breath Standard and complete this form.

Chemical Test Operator Name: <u>William R Petty</u>		
Location: <u>GF AF</u>		
Intoxilyzer® Serial Number: <u>80-004949</u>		
Gas Lot Number: <u>659358</u>	Gas Cylinder Number: <u>39</u>	Expiration Date: <u>09/03/2011</u>
<input type="checkbox"/> New Cylinder Installed <input checked="" type="checkbox"/> Cylinder Not Changed		
Instrument Test Results (Report to 3 Digits; Example, 0.081%) Test 1: <u>0.076</u> % Test 2: <u>0.677</u> % Test 3: <u>0.077</u> %		
(ATTACH TEST RECORD)		

Will. Petty

Chemical Test Operator Signature

16 Aug 2010

Date

Deb Kashur

Reviewed By

27 Oct 2010

Date

80-004949  
P. 57 of 58  
RM

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004949  
Location = GFAP      8164.13.00 06/09  
08/16/2010      08:23

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	08:25
02 Std. Gas	0.076	08:25
03 Room Air	0.000	08:26
04 Std. Gas	0.077	08:26
05 Room Air	0.000	08:27
06 Std. Gas	0.077	08:27
07 Room Air	0.000	08:28

Lot No = 659358

Cyl No = 39

Exp Date = 09/03/2011

County = 18

Oper No. = 120011



Operator Signature  
WILLIAM R. PETTY

Remarks:

Form 106-I8000

80-004949  
P. 58 of 58  
WLP