



ETHANOL BREATH STANDARD CYLINDER REPORT
 OFFICE OF ATTORNEY GENERAL
 CRIME LABORATORY DIVISION
 Toxicology Section/Breath Alcohol Program
 SFN 59282 (1-2010)

Only Chemical Test Operators may change the Ethanol Breath Standard and complete this form.

Chemical Test Operator Name: <u>Joseph Knowski</u>		
Location: <u>Langdon SO. Cavalier Co.</u>		
Intoxilyzer® Serial Number: <u>80-004937</u>		
Gas Lot Number: <u>659358</u>	Gas Cylinder Number: <u>22</u>	Expiration Date: <u>9-3-2011.</u>
<input type="checkbox"/> New Cylinder Installed <input checked="" type="checkbox"/> Cylinder Not Changed		
Instrument Test Results (Report to 3 Digits; Example, 0.081%) Test 1: <u>0.077</u> % Test 2: <u>0.077</u> % Test 3: <u>0.076</u> %		
(ATTACH TEST RECORD)		

Joseph Knowski
 Chemical Test Operator Signature

Deb Kashur
 Reviewed By

8-17-2010
 Date

23 Aug 2010
 Date


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004937
Location = LANG 8164.13.00 06/09
08/17/2010 13:26

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:27
02 Std. Gas	0.077	13:28
03 Room Air	0.000	13:28
04 Std. Gas	0.077	13:29
05 Room Air	0.000	13:29
06 Std. Gas	0.076	13:29
07 Room Air	0.000	13:30

Lot No = 659358
Cyl No = 22
Exp Date = 09/03/2011
County = 10 Oper No. = 109208



Operator Signature
JOSEPH J. KNOWSKI

Remarks:

Form 106-18000

80-004937
P. 45 of 45
JAM