



ETHANOL BREATH STANDARD CYLINDER REPORT

OFFICE OF ATTORNEY GENERAL
CRIME LABORATORY DIVISION
Toxicology Section/Breath Alcohol Program
SFN 59282 (1-2010)

Only Chemical Test Operators may change the Ethanol Breath Standard and complete this form.

Chemical Test Operator Name: <u>Richard D. Johnson</u>		
Location: <u>McLean L.E.C. Washburn ND</u>		
Intoxilyzer® Serial Number: <u>80-004196</u>		
Gas Lot Number: <u>659357</u>	Gas Cylinder Number: <u>47</u>	Expiration Date: <u>09/03/11</u>
<input type="checkbox"/> New Cylinder Installed	<input checked="" type="checkbox"/> Cylinder Not Changed	
Instrument Test Results (Report to 3 Digits; Example, 0.081%)		
Test 1: <u>0.080</u> %		
Test 2: <u>0.080</u> %		
Test 3: <u>0.080</u> %		
(ATTACH TEST RECORD)		

Richard D. Johnson
Chemical Test Operator Signature

8-18-10
Date

Deb Kashur
Reviewed By

23 Aug 2010
Date

80-004196
P. 91 of 92
xcm


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004196
Location = WASH 8164.13.00 06/09
08/18/2010 11:27

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:27
02 Std. Gas	0.080	11:28
03 Room Air	0.000	11:28
04 Std. Gas	0.080	11:29
05 Room Air	0.000	11:29
06 Std. Gas	0.080	11:30
07 Room Air	0.000	11:30

Lot No = 659357
Cyl No = 47
Exp Date = 09/03/2011
County = 28 Oper No. = 118707


Operator Signature
RICHARD JOHNSON

Remarks:

Form 106-I8000

80-004196
P. 92 of 92
Ym