



OFFICE OF ATTORNEY GENERAL
 Crime Laboratory Division
 2641 East Main Avenue
 P.O. Box 937
 Bismarck, North Dakota 58502-0937

TEL 701-328-6159

ETHANOL BREATH STANDARD CYLINDER INSTALLATION FORM

(Only Chemical Test Operators may change the Ethanol Breath Standard and complete this form.)

Chemical Test Operator Name: JAY GALLAGHER

Location: WALSH LEC, GRAFTON, ND

Intoxilyzer Serial No.: ~~659358~~ 80-003072

Lot Number: 659358

Cylinder Number: 16

Expiration Date: 9-3-11

Instrument Test Results (Report to 3 Digits; Example, 0.081%)

Test 1: 0. 081 %

Test 2: 0. 080 %

Test 3: 0. 080 %

(ATTACH TEST RECORD)

J. Gallagher
 Chemical Test Operator Signature

8-20-2010
 Date Cylinder Changed CHECKED

Deb Kashner
 Reviewed By

DK
31 Aug 2010 07 Sept 2010
 Date

Form 105-G
 Toxicology Laboratory
 Crime Laboratory Division

80-003072
 P. 93 of 95
 JH

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-003072
Location = GRAF 8164.13.00 06/09
08/20/2010 18:27

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	18:28
02 Std. Gas	0.081	18:28
03 Room Air	0.000	18:29
04 Std. Gas	0.080	18:29
05 Room Air	0.000	18:30
06 Std. Gas	0.080	18:30
07 Room Air	0.000	18:31

Lot No = 659358
Cyl No = 16
Exp Date = 09/03/2011
County = 50 Oper No. = 120806

Operator Signature
JAMES J. GALLAGHER

Remarks:

Form 106-I8000

80-003072
P. 94 of 95
JM

.6_AF63iGIAAS%2FoIHO... 9/4/2010

dbb KASIKUR

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SW 80-303072
Location = GRAF 8164.13.00 06/09
08/20/2010 18:27

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	18:28
02 Std. Gas	0.081	18:28
03 Room Air	0.000	18:29
04 Std. Gas	0.080	18:29
05 Room Air	0.000	18:30
06 Std. Gas	0.080	18:30
07 Room Air	0.000	18:31

Lot No = 659358
Cyl No = 16
Exp Date = 09/03/2011
County = 50
Oper No. = 120806

[Signature]
Operator Signature
JAMES J. GALLAGHER

Remarks:

80-003072
P. 95 of 95
JMG