


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-003059
Location = WFPD 8164.13.00 06/09
08/23/2010 15:02

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	15:03
02 Std. Gas	0.080	15:03
03 Room Air	0.000	15:04
04 Std. Gas	0.079	15:04
05 Room Air	0.000	15:05
06 Std. Gas	0.079	15:05
07 Room Air	0.000	15:06

Lot No = 659357 *DK 29 April*
Cyl No = 10
Exp Date = ~~12/23/2011~~ *09/03/11*
County = 09 Oper No. = 039617



Operator Signature
PETER D. NIELSEN

Remarks:

Form 106-I8000

80-003059
P. 107 of 107
km
80-003059
P. 96 of 96
km