



ETHANOL BREATH STANDARD CYLINDER REPORT
 OFFICE OF ATTORNEY GENERAL
 CRIME LABORATORY DIVISION
 Toxicology Section/Breath Alcohol Program
 SFN 59282 (1-2010)

Only Chemical Test Operators may change the Ethanol Breath Standard and complete this form.

Chemical Test Operator Name: <u>Shannon P. Liebald</u>		
Location: <u>Mercer County Sheriff's Dept.</u>		
Intoxilyzer® Serial Number: <u>80-003060</u>		
Gas Lot Number: <u>659357</u>	Gas Cylinder Number: <u>43</u>	Expiration Date: <u>09-03-11</u>
<input type="checkbox"/> New Cylinder Installed	<input checked="" type="checkbox"/> Cylinder Not Changed	-
Instrument Test Results (Report to 3 Digits; Example, 0.081%)		
Test 1: 0. <u>080</u> %		
Test 2: 0. <u>080</u> %		
Test 3: 0. <u>079</u> %		
(ATTACH TEST RECORD)		

Shannon P. Liebald
 Chemical Test Operator Signature

8-24-10
 Date

Deb Kashner
 Reviewed By

31 Aug 2010
 Date

80-003060
 P. 91 of 92
 YH


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-003060
Location = STAT 8164.13.00 06/09
08/24/2010 01:41

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	01:42
02 Std. Gas	0.080	01:42
03 Room Air	0.000	01:43
04 Std. Gas	0.080	01:43
05 Room Air	0.000	01:44
06 Std. Gas	0.079	01:44
07 Room Air	0.000	01:45

Lot No = 659357
Cyl No = 43
Exp Date = 09/03/2011
County = 29 Oper No. = 020904



Operator Signature
SHANNON P. LIEBOLD

Remarks:

Form 106-I8000

80-003060
P. 92 of 92
JFM