



INTOXILYZER® 8000 INSTALLATION AND REPAIR CHECKOUT
 OFFICE OF ATTORNEY GENERAL
 CRIME LABORATORY DIVISION
 Toxicology Section/Breath Alcohol Program
 SFN 59281 (03-2014)

Serial Number 80-005953	Instrument Location Ellendale
Reason for Install/Repair <input checked="" type="checkbox"/> Install After Receiving From Crime Laboratory <input type="checkbox"/> Install After Location Change <input type="checkbox"/> Other (Specify) _____	

Check When Done:

- 1. Surge Protector Installed/Properly Grounded.
- 2. Telephone Line Connected to Intoxilyzer® 8000.
- 3. Breath Tube Heated.
- 4. Enter Preliminary Data (i.e. Date, Time, DST (Y), and Location; Level 2, Function E).
- 5. Scan/Enter Gas Cylinder Information (Level 1, Function S).
- 6. Run Tests:
 - A. Print Test (Level 1, Function P).
 - B. ACA Test (Level 1, Function C).
 - C. RFI Test (CMS Mode or Level 1, Function B or C; Key Radio During Test).
- 7. Repair and/or Maintenance Performed (if Any): n/a
- 8. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.
- 9. File Previous Intoxilyzer® Record (SFN50496, Form 120-G) at the Intoxilyzer® Location at the Agency.
- 10. Send the Following to the Crime Laboratory: Completed Intoxilyzer®8000 Installation and Repair Checkout (SFN59281, Form 104-G), Print Test, ACA Test, and RFI Test.

Field Inspector Signature <i>Chris Ed</i>	Date 8-24-16
Reviewed By (Crime Laboratory Use Only) <i>Deb Shanaver</i>	Date 25 Aug 16

This record on file at the Office of Attorney General, Crime Laboratory Division, in the County of Burleigh, North Dakota, is certified to be a true and correct copy of the documents received.

Certified By: *Charles E. Ed* Certified Date: 21 Nov. 16

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005953
Location = DCSO 8164.13.00 06/09
08/24/2016 09:09

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	09:10
02 Std. Gas	0.082	09:10
03 Room Air	0.000	09:11
04 Std. Gas	0.081	09:11
05 Room Air	0.000	09:11
06 Std. Gas	0.081	09:12
07 Room Air	0.000	09:12

Lot No = 09216080A3
Cyl No = 36
Exp Date = 05/05/2018
County = 11 Oper No. = 020504



Operator Signature
CHRIS ESTES

Remarks:

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005953
Location = DCSO 8164.13.00 06/09
08/24/2016 09:05

DRY CAL CHECK

Test	AC	Time
01 Room Air	RFI*	09:06
02 Room Air	0.000	09:06

*Invalid Test

Inhibited (RFI)

Lot No = 09216080A3

Cyl No = 36

Exp Date = 05/05/2018

County = 11

Oper No. = 020504



Operator Signature
CHRIS ESTES

Remarks: