



# ETHANOL BREATH STANDARD CYLINDER REPORT

OFFICE OF ATTORNEY GENERAL  
CRIME LABORATORY DIVISION  
Toxicology Section/Breath Alcohol Program  
SFN 59282 (1-2010)

Only Chemical Test Operators may change the Ethanol Breath Standard and complete this form.

Chemical Test Operator Name: <u>Phil Wolf</u>		
Location: <u>Cavalier, ND</u>		
Intoxilyzer® Serial Number: <u>80-002669</u>		
Gas Lot Number: <u>659357</u>	Gas Cylinder Number: <u>9</u>	Expiration Date: <u>9/3/11</u>
<input type="checkbox"/> New Cylinder Installed	<input type="checkbox"/> Cylinder Not Changed	
Instrument Test Results (Report to 3 Digits; Example, 0.081%)		
Test 1: 0. <u>076</u> %		
Test 2: 0. <u>076</u> %		
Test 3: 0. <u>070</u> %		
(ATTACH TEST RECORD)		

P. Wolf  
Chemical Test Operator Signature

8/26/10  
Date

Deb Kashner  
Reviewed By

31 Aug 2010  
Date

80-002669  
P. 96 of 97  
AM


Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-002669  
Location = CAVA      8164.13.00 06/09  
08/26/2010      07:54

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	07:54
02 Std. Gas	0.076	07:55
03 Room Air	0.000	07:55
04 Std. Gas	0.076	07:55
05 Room Air	0.000	07:56
06 Std. Gas	0.077	07:56
07 Room Air	0.000	07:57

Lot No = 659357  
Cyl No = 9  
Exp Date = 09/03/2011  
County = 34      Oper No. = 120547

  
\_\_\_\_\_  
Operator Signature  
PHIL WOLF

Remarks:

Form 106-I8000

80-002669  
P. 97 of 97  
FM