



**ETHANOL BREATH STANDARD CYLINDER REPORT**  
OFFICE OF ATTORNEY GENERAL  
CRIME LABORATORY DIVISION  
Toxicology Section/Breath Alcohol Program  
SFN 59282 (1-2010)

Only Chemical Test Operators may change the Ethanol Breath Standard and complete this form.

Chemical Test Operator Name: <u>Duisen, SHANNON P</u>		
Location: <u>TIOGA P.D</u>		
Intoxilyzer® Serial Number: <u>80-004938</u>		
Gas Lot Number: <u>709352</u>	Gas Cylinder Number: <u>21</u>	Expiration Date: <u>12/23/11</u>
<input type="checkbox"/> New Cylinder Installed	<input checked="" type="checkbox"/> Cylinder Not Changed	
Instrument Test Results (Report to 3 Digits; Example, 0.081%)		
Test 1: <u>0.077</u> %		
Test 2: <u>0.079</u> %		
Test 3: <u>0.078</u> %		
(ATTACH TEST RECORD)		

Chemical Test Operator Signature

8/28/10  
Date

Deb Keshur  
Reviewed By

07 Sept '10  
Date

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004938  
Location = TIOG      8164.13.00 06/09  
08/28/2010      16:24

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	16:26
02 Std. Gas	0.077	16:27
03 Room Air	0.000	16:27
04 Std. Gas	0.079	16:27
05 Room Air	0.000	16:28
06 Std. Gas	0.078	16:28
07 Room Air	0.000	16:29

Lot No = 709352  
Cyl No = 21  
Exp Date = 12/23/2011  
County = 53      Oper No. = 020632



Operator Signature  
SHANNON DUISEN

Remarks:

Form 106-I8000

80-004938  
p. 54 of 54  
Jm