



**ETHANOL BREATH STANDARD CYLINDER REPORT**  
 OFFICE OF ATTORNEY GENERAL  
 CRIME LABORATORY DIVISION  
 Toxicology Section/Breath Alcohol Program  
 SFN 59282 (1-2010)

Only Chemical Test Operators may change the Ethanol Breath Standard and complete this form.

Chemical Test Operator Name: <u>David Feland</u>		
Location: <u>Kidder S.O.</u>		
Intoxilyzer® Serial Number: <u>80-004186</u>		
Gas Lot Number: <u>659357</u>	Gas Cylinder Number: <u>37</u>	Expiration Date: <u>09/03/11</u>
<input type="checkbox"/> New Cylinder Installed <input checked="" type="checkbox"/> Cylinder Not Changed		
Instrument Test Results (Report to 3 Digits; Example, 0.081%) Test 1: <u>0.080</u> % Test 2: <u>0.079</u> % Test 3: <u>0.080</u> %		
(ATTACH TEST RECORD)		

David Feland  
 Chemical Test Operator Signature

31 Aug 10  
 Date

Deb Kashur  
 Reviewed By

07 Sept 2010  
 Date


Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004186  
Location = STEE      8164.13.00 06/09  
08/31/2010      14:14

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:15
02 Std. Gas	0.080	14:15
03 Room Air	0.000	14:16
04 Std. Gas	0.079	14:16
05 Room Air	0.000	14:16
06 Std. Gas	0.080	14:17
07 Room Air	0.000	14:17

Lot No = 659357  
Cyl No = 37  
Exp Date = 09/03/2011  
County = 22      Oper No. = 119404



Operator Signature  
DAVID R. FELAND

Remarks:

Form 106-I8000

80-004186  
P. 88 of 88 *am*