

✓ CML

INTOXILYZER 5000 G and 8000 INSPECTION

Serial No.: 80-004200

Location: TOXL

Check When Done

- A. Is the warm-up time less than 20 minutes? ✓
- B. Is a three-pronged grounded outlet used? ✓
- C. Is the breath tube heated? ✓
- D. Is the diagnostic check complete? ✓
- E. Is the time, date, and year correct (re-set if necessary)? ✓
- F. Print test. (Attach test record.) ✓
- G. Low AC. Use <0.03 AC in ACA mode. (Attach test record.) ✓
Sim SN DR5190 Lot No. — Rep AC 0.021
- H. Linearity Test. Use ≥ 0.25 AC in ACA mode. (Attach test record.) ✓
Sim SN DR5113 Lot No. — Rep AC 0.27 AC
- I. Interferent check. Use 0.05% acetone plus 0.10 AC ethanol in ABA mode. ✓
(Attach test record.)
Sim SN DR5135 Lot No. — Rep AC 0.10 AC ET ACETONE
- J. RFI check. Run CMS mode. Key radio on first room air. ✓
(Attach test record.)
- K. Calibration Check. Use Ethanol Breath Standard cylinder. Do three sets in ACA mode.
(Attach test record.)

Lot No. 617774 Cyl No. 12 Exp. Date 5/28/2011
Test 1 0.080 Test 4 0.080 Test 7 0.079
Test 2 0.079 Test 5 0.079 Test 8 0.079
Test 3 0.080 Test 6 0.079 Test 9 0.079
Average 0.079 AC

L. Remarks/Maintenance record: Annual Inspection
1 September 2009 Mary Beard
Date Field Inspector's Signature
04 Sept 2009 Deb Kashner
Date Reviewed by

08/09

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004200
Location = TOXL 8164.12.00 06/09
09/01/2009 14:27

***** Printer Test *****

abcdefghijklmnopqrstuvwxy1234567890-=_|
ABCDEFGHIJKLMNPOQRSTUVWXYZ!@#\$\$%^&*()_+?

abcdefghijklmnopqrstuvwxy1234567890-=_|
ABCDEFGHIJKLMNPOQRSTUVWXYZ!@#\$\$%^&*()_+?

Current Instrument Setup


Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? Yes
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 617774
Standard Cyl #? 12
Standard Expiration? 05/28/2011
Oper No? 107501

Flow Cal. Date: 06/04/2009
Slope 715
Intercept -655593

IR Calibration Date: 06/03/2009
 3um 9um

0th Coef(*100): -17852 -24276
1st Coef(*100): 261183 140006
2nd Coef(*100): 2285 945
H2O adj(mg/l*10k): 600 514

***** Printer Test End *****



Operator Signature
MARGY PEARSON

Remarks:

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004200
Location = TOXL 8164.12.00 06/09
09/01/2009 16:17

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	16:18
02 Std. Sol.	0.017	16:18
03 Room Air	0.000	16:19
04 Std. Sol.	0.018	16:20
05 Room Air	0.000	16:20
06 Std. Sol.	0.018	16:21
07 Room Air	0.000	16:21


08 Sim Temp = 34.0°C

Simul Ser No = DR0000

Std Sol No = 0000

County = 08

Oper No. = 107501



Operator Signature
MARGY PEARSON

Remarks:

Low AC 0.021

Form 106-I8000

*80-004200
P. 3 of 9 FM*

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004200
Location = TOXL 8164.12.00 06/09
09/01/2009 17:09

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	17:09
02 Std. Sol.	0.289	17:10
03 Room Air	0.000	17:11
04 Std. Sol.	0.292	17:11
05 Room Air	0.000	17:12
06 Std. Sol.	0.290	17:13
07 Room Air	0.000	17:13


08 Sim Temp = 34.0°C

Simul Ser No = DR0000

Std Sol No = 0000

County = 08

Oper No. = 107501



Operator Signature
MARGY PEARSON

Remarks:

Linearity

Form 106-I8000

80-004200
p. 4 of 9 FA

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004200
Location = TOXL 8164.12.00 06/09
09/01/2009 17:03

Test	AC	Time
01 Room Air	0.000	17:04
02 *Subject Test	INT*	17:04
03 Room Air	0.000	17:05

*Invalid Test
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 11/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = NA
Dr. Lic. = ND/DIS821456
Simul Ser No = DR0000
Std Sol No = 0000
County = 08 Oper No. = 107501



Operator Signature
MARGY PEARSON

Remarks:

Interferent Alc + Acetone

Form 106-I8000

*80-004200
P. 5 of 9 JAH*


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004200
Location = TOXL 8164.12.00 06/09
09/01/2009 15:48

Test	AC	Time
01 Room Air	RFI*	15:48
02 Room Air	0.000	15:49

*Invalid Test
 Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 11/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = NA
Dr. Lic. = ND/DIS821456
Simul Ser No = DR0000
Std Sol No = 0000
County = 08 Oper No. = 107501



Operator Signature
MARGY PEARSON

Remarks:

RFI

Form 106-I8000

80-004200
P. 6/9 KH


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004200
Location = TOXL 8164.12.00 06/09
09/01/2009 13:39

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:40
02 Std. Gas	0.080	13:40
03 Room Air	0.000	13:40
04 Std. Gas	0.079	13:41
05 Room Air	0.000	13:41
06 Std. Gas	0.080	13:42
07 Room Air	0.000	13:42

Lot No = 617774
Cyl No = 12
Exp Date = 05/28/2011
County = 08 Oper No. = 107501



Operator Signature
MARGY PEARSON

Remarks:

ACA

Form 106-I8000

80-004200
P. 2-4 7 FH


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004200
Location = TOXL 8164.12.00 06/09
09/01/2009 13:43

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:43
02 Std. Gas	0.080	13:44
03 Room Air	0.000	13:44
04 Std. Gas	0.079	13:45
05 Room Air	0.000	13:45
06 Std. Gas	0.079	13:45
07 Room Air	0.000	13:46

Lot No = 617774
Cyl No = 12
Exp Date = 05/28/2011
County = 08 Oper No. = 107501



Operator Signature
MARGY PEARSON

Remarks:

ACA

Form 106-I8000

80-004200
P. 8-19-09


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004200
Location = TOXL 8164.12.00 06/09
09/01/2009 13:51

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:52
02 Std. Gas	0.079	13:52
03 Room Air	0.000	13:52
04 Std. Gas	0.079	13:53
05 Room Air	0.000	13:53
06 Std. Gas	0.079	13:54
07 Room Air	0.000	13:54

Lot No = 617774
Cyl No = 12
Exp Date = 05/28/2011
County = 08 Oper No. = 107501



Operator Signature
MARGY PEARSON

Remarks:

ACA

Form 106-I8000

80-004200
R 9 of 9 FM