



ETHANOL BREATH STANDARD CYLINDER REPORT

OFFICE OF ATTORNEY GENERAL
CRIME LABORATORY DIVISION
Toxicology Section/Breath Alcohol Program
SFN 59282 (1-2010)

Only Chemical Test Operators may change the Ethanol Breath Standard and complete this form.

Chemical Test Operator Name: <u>Jonathan M Toffeland</u>		
Location: <u>MOHL</u>		
Intoxilyzer® Serial Number: <u>80-004198</u>		
Gas Lot Number: <u>659358</u>	Gas Cylinder Number: <u>27</u>	Expiration Date: <u>9/3/11</u>
<input type="checkbox"/> New Cylinder Installed	<input checked="" type="checkbox"/> Cylinder Not Changed	
Instrument Test Results (Report to 3 Digits; Example, 0.081%)		
Test 1: <u>0.077</u> %		
Test 2: <u>0.078</u> %		
Test 3: <u>0.078</u> %		
(ATTACH TEST RECORD)		

Jonathan M Toffeland
Chemical Test Operator Signature

9-1-10
Date

Deb Keshum
Reviewed By

07 Sept 2010
Date


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004198
Location = MOHL 8164.13.00 06/09
09/01/2010 09:54

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	09:55
02 Std. Gas	0.077	09:55
03 Room Air	0.000	09:55
04 Std. Gas	0.078	09:56
05 Room Air	0.000	09:56
06 Std. Gas	0.078	09:57
07 Room Air	0.000	09:57

Lot No = 659358
Cyl No = 27
Exp Date = 09/03/2011
County = 38 Oper No. = 100915



Operator Signature
JONATHAN M. TOFTELAND

Remarks:

Form 106-I8000

80-004198
P. 100 of 100
JTM