



OFFICE OF ATTORNEY GENERAL  
 CRIME LABORATORY DIVISION  
 Toxicology Section/Breath Alcohol Program  
 Instrument Certification and Maintenance Unit

BrF-005-8

Q. Flow Monitor. Level 3 Func D Sub F. R 164 F 0,000 L/s

R. Temperature Monitor. Level 3 Func D Sub T  
 C 47.0 A 44.5 I 0.0  
 B 45.0 V 0.0

S. Gas Tank Monitor. Level 3 Func D Sub G  
 T. Configure Wet Bath Sim. Level 1 Func S  
504 psi Intox 500 psi Reg

U. Zero Alcohol Test Level 1 Func C  
DR 3409 Sim SN — Lot 0,000 Rep Value  
0.000 AC 0.000 AC 0.000 AC

V. Low AC Test. Level 1 Func C  
DR 5144 Sim SN — Lot 0.021 Rep Value  
0.018 AC 0.019 AC 0.019 AC

W. Linearity Test. Level 1 Func C  
DR 5187 Sim SN — Lot 0.290 Rep Value  
0.287 AC 0.288 AC 0.290 AC

X. Interferent Test.  
 Level 1 Func B  
 Acetone DR 3845 Sim SN — Lot 0.0590 Conc  
 Display  Printout  
 Methanol G 10573 Sim SN — Lot 0.1090 Conc  
 Display  Printout  
 IPA DR 5191 Sim SN — Lot 0.1090 Conc  
 Display  Printout

Y. Ambient Test  
 Display  Printout

Z. Invalid Sample Test  
 Display  Printout

AA. RFI Detector Test. CMS Mode  
 Display  Printout

BB. Configure Dry Gas Standard. Level 1 Func S

CC. Stability Test at 0.080AC. Level 3 Func T Sub S  
0.0790 Ave. 0.0003 Std Dev 0.42% Rel Std Dev

DD. Maintenance/Repairs Initial inspection

02 Sept 09  
 Date  
4 Sept 09  
 Date

Deb Kashner  
 Signature  
Mary Pearson  
 Reviewed by

P. 2 of 17  
 FM

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000        SN 80-002667  
Location = TOXL                8164.12.00 06/09  
09/02/2009                      08:23

\*\*\*\*\* Printer Test \*\*\*\*\*

abcdefghijklmnopqrstuvwxyz1234567890-=\_|  
ABCDEFGHIJKLMNPOQRSTUVWXYZ!@#%^&\*()\_+?

abcdefghijklmnopqrstuvwxyz1234567890-=\_|  
ABCDEFGHIJKLMNPOQRSTUVWXYZ!@#%^&\*()\_+?

Current Instrument Setup

Data Entry Mode:            Enabled  
Start Test Sequence:        DABACABA  
Display Prelim Rslt?        Yes  
Display Third Digit?        Yes  
Inhib Printer(Y/N)?        No  
Display Volume?            No  
Disable On Memfull?        No  
# of Print Copies?         1  
Select Std (D/W/I)?        Dry  
Standard Value?            0.080  
Standard Lot #?            617774  
Standard Cyl #?            38  
Standard Expiration?       05/28/2011  
Oper No?                    777777

Flow Cal. Date:            09/10/2007  
Slope                      693  
Intercept                  -570251

IR Calibration Date:        09/10/2007

	3um	9um
0th Coef(*100):	-21400	-13991
1st Coef(*100):	268177	135590
2nd Coef(*100):	1202	658
H2O adj(mg/l*10k):	369	165

\*\*\*\*\* Printer Test End \*\*\*\*\*

Deb Kashur

Operator Signature  
DEB KASHUR

Remarks:

Print Test

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-002667  
Location = TOXL      8164.12.00 06/09  
09/01/2009      16:33

\*\*\*\*\* Printer Test \*\*\*\*\*

12 char/line  
abcdefghijkl  
ABCDEFGHIJKL

16 char per line  
abcdefghijklmnop  
ABCDEFGHIJKLMNPO

21 chars per line  
abcdefghijklmnopqrstu  
ABCDEFGHIJKLMNOPQRSTU

24 characters per line  
abcdefghijklmnopqrstuvw  
ABCDEFGHIJKLMNOPQRSTUWX

32 characters per line  
abcdefghijklmnopqrstuvwxy123456  
ABCDEFGHIJKLMNOPQRSTUWXYZ123456

42 characters per line  
abcdefghijklmnopqrstuvwxy1234567890abcdef  
ABCDEFGHIJKLMNOPQRSTUWXYZ1234567890ABCDEF

Current Instrument Setup

Data Entry Mode: Enabled  
Start Test Sequence: DABACABA  
Display Prelim Rslt? Yes  
Display Third Digit? Yes  
Inhib Printer(Y/N)? No  
Display Volume? No  
Disable On MemFull? No  
# of Print Copies? 1  
Select Std (D/W/I)? Wet  
Standard Value? 0.110  
Standard Solution #: 555  
Sim Serial #: DR444444  
Oper No? 77777

Flow Cal. Date: 09/10/2007  
Slope: 693  
Intercept: -570251

IR Calibration Date: 09/10/2007

	Sum	Sum
0th Coef(*100):	-21405	-13991
1st Coef(*100):	268177	135550
2nd Coef(*100):	1202	658
H2O adj(mg/l*10K):	369	165

\*\*\*\*\* Test End \*\*\*\*\*

*Deb Kasher*

Operator Signature  
DEB KASHUR

Remarks:

*Print Test*

Form 106-18000

*P. 4 of 12*

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-002667  
Location = TOXL      8164.12.00 06/09  
09/02/2009      08:49

Test	AC	Time
01 Room Air	0.000	08:49
02 *Subject Test	REF*	08:50
03 Room Air	0.000	08:50

\*Subject Test Refused

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 02/01/1992  
Sub Sex = Female      Weight = 150  
Test = DUI      Cit = NA  
Dr. Lic. = ND/DIS921456  
Lot No = 617774  
Cyl No = 38  
Expiration Date = 05/28/2011  
County = 08      Oper No. = 777777

Deb Kashur

Operator Signature  
DEB KASHUR

Remarks:

Camera / "R" check

Form 106-I8000

P. 5 of 17  
PM

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-002667  
Location = TOXL      8164.12.00 06/09  
09/02/2009      08:51

DIAGNOSTIC

Voltage/Current Test	Pass
RAM Test	Pass
EEPROM Checksum Test	Pass
Real Time Clock Test	Pass
DSP Test	Pass
Analytical Stability Test	Pass
Modem Test	Pass
Temp Regulation Test	Pass

Oper No. = 777777

*Deb Kasher*

Operator Signature  
DEB KASHUR

Remarks:

*Diagnostic test*

Form 106-I8000

*P. Coft/7  
FM*

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-002667  
Location = TOXL      8164.12.00 06/09  
09/02/2009      10:22

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:23
02 Std. Sol.	0.000	10:25
03 Room Air	0.000	10:25
04 Std. Sol.	0.000	10:26
05 Room Air	0.000	10:26
06 Std. Sol.	0.000	10:27
07 Room Air	0.000	10:27

08 Sim Temp = 34.0°C

Simul Ser No = DR0000  
Std Sol No = 000  
County = 08      Oper No. = 777777

Deb Kasher  
Operator Signature  
DEB KASHUR

Remarks:

*zero alcohol*

Form 106-I8000

*P. 7/17  
FM*

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-002667  
Location = TOXL      8164.12.00 06/09  
09/02/2009      13:38

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:39
02 Std. Sol.	0.018	13:46
03 Room Air	0.000	13:46
04 Std. Sol.	0.019	13:47
05 Room Air	0.000	13:47
06 Std. Sol.	0.019	13:48
07 Room Air	0.000	13:49

08 Sim Temp = 34.0°C

Simul Ser No = DR0000

Std Sol No = 000

County = 08

Oper No. = 777777

*Deb Kashur*

Operator Signature  
DEB KASHUR

Remarks:

*Low AC*

Form 106-I8000

*P. 8/17  
JFM*

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-002667  
Location = TOXL      8164.12.00 06/09  
09/02/2009      10:43

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:44
02 Std. Sol.	0.287	10:46
03 Room Air	0.000	10:46
04 Std. Sol.	0.288	10:47
05 Room Air	0.000	10:47
06 Std. Sol.	0.290	10:48
07 Room Air	0.000	10:49

08 Sim Temp = 34.0°C

Simul Ser No = DR0000  
Std Sol No = 000  
County = 08      Oper No. = 777777

  
\_\_\_\_\_  
Operator Signature  
DEB KASHUR

Remarks:  
*Linearity test*

Form 106-18000

*P. 9 of 17  
FM*

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-002667  
Location = TOXL      8164.12.00 06/09  
09/02/2009      10:30

Test	AC	Time
01 Room Air	0.000	10:30
02 *Subject Test	INT*	10:31
03 Room Air	0.000	10:31

\*Invalid Test  
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 02/01/1992  
Sub Sex = Female      Weight = 150  
Test = DUI      Cit = NA  
Dr. Lic. = ND/DIS921456  
Simul Ser No = DR0000  
Std Sol No = 000  
County = 08      Oper No. = 777777

Deb Kashur

Operator Signature  
DEB KASHUR

Remarks:

*acetone interferent*

Form 106-I8000

*P. 10/17  
PM*

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000        SN 80-002667  
Location = TOXL                8164.12.00 06/09  
09/02/2009                      13:55

Test	AC	Time
01 Room Air	0.000	13:56
02 *Subject Test	INT*	13:57
03 Room Air	0.000	13:57

\*Invalid Test  
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 02/01/1992  
Sub Sex = Female                Weight = 150  
Test = DUI                        Cit = NA  
Dr. Lic. = ND/DIS921456  
Simul Ser No = DR0000  
Std Sol No = 000  
County = 08                      Oper No. = 777777

Deb Kasher  
Operator Signature  
DEB KASHUR

Remarks:

*Methanol interferent*

Form 106-I8000

*P. 11 Oct 17  
jrm*

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000        SN 80-002667  
Location = TOXL                    8164.12.00 06/09  
09/02/2009                            14:02

Test	AC	Time
01 Room Air	0.000	14:03
02 *Subject Test	INT*	14:03
03 Room Air	0.000	14:04

\*Invalid Test  
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 02/01/1992  
Sub Sex = Female                    Weight = 150  
Test = DUI                            Cit = NA  
Dr. Lic. = ND/DIS921456  
Simul Ser No = DR0000  
Std Sol No = 000  
County = 08                            Oper No. = 777777

*Deb Kashur*

Operator Signature  
DEB KASHUR

Remarks:

*IPA interferent*

Form 106-I8000

*P. 12 of 17  
JKM*

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000            SN 80-002667  
Location = TOXL                    8164.12.00 06/09  
09/02/2009                            14:55

Test	AC	Time
01 Room Air	AMB*	14:55
02 Room Air	0.000	14:56

\*Invalid Test  
Check Ambient Conditions

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 02/01/1992  
Sub Sex = Female                    Weight = 150  
Test = DUI                            Cit = NA  
Dr. Lic. = ND/DIS921456  
Simul Ser No = DR0000  
Std Sol No = 000  
County = 08                            Oper No. = 777777

Deb Kasher  
Operator Signature  
DEB KASHUR

Remarks:

*Ambient fail test*

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000        SN 80-002667  
Location = TOXL                    8164.12.00 06/09  
09/02/2009                            16:22

Test	AC	Time
01 Room Air	0.000	16:23
02 Invalid Sample	X.XXX	16:23
03 Room Air	PUR*	16:24

\*Purge Fail  
\*\*Invalid Test - Mouth Alcohol

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 02/01/1992  
Sub Sex = Female                    Weight = 150  
Test = DUI                            Cit = NA  
Dr. Lic. = ND/DIS921456  
Simul Ser No = DR0000  
Std Sol No = 000  
County = 08                            Oper No. = 777777

Deb Kashur

Operator Signature  
DEB KASHUR

Remarks:

*no sample*  
*Invalid test*

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000        SN 80-002667  
Location = TOXL                8164.12.00 06/09  
09/02/2009                      14:45

Test	AC	Time
01 Diagnostic	OK	14:46
02 Room Air	RFI*	14:46
03 Room Air	0.000	14:47

\*Invalid Test  
Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 02/01/1992  
Sub Sex = Female                Weight = 150  
Test = DUI                        Cit = NA  
Dr. Lic. = ND/DIS921456  
Simul Ser No = DR0000  
Std Sol No = 000  
County = 08                      Oper No. = 777777

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.

Deb Kasher

Operator Signature  
DEB KASHUR

Remarks:

RFI test

Form 106-I8000

P.15 of 17  
FK

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000        SN 80-002667  
Location = TOXL                8164.12.00 06/09  
09/01/2009                      16:35

Test	AC	Time
01 Room Air	0.000	16:36
02 Std. Gas	0.079	16:36
03 Room Air	0.000	16:36
04 Std. Gas	0.079	16:37
05 Room Air	0.000	16:37
06 Std. Gas	0.078	16:38
07 Room Air	0.000	16:38
08 Std. Gas	0.079	16:38
09 Room Air	0.000	16:39
10 Std. Gas	0.079	16:39
11 Room Air	0.000	16:40
12 Std. Gas	0.079	16:40
13 Room Air	0.000	16:41
14 Std. Gas	0.079	16:41
15 Room Air	0.000	16:41
16 Std. Gas	0.079	16:42
17 Room Air	0.000	16:42
18 Std. Gas	0.079	16:43
19 Room Air	0.000	16:43
20 Std. Gas	0.079	16:44
21 Room Air	0.000	16:44
22 Std. Gas	0.079	16:44
23 Room Air	0.000	16:45
24 Std. Gas	0.079	16:45
25 Room Air	0.000	16:46
26 Std. Gas	0.079	16:46
27 Room Air	0.000	16:47
28 Std. Gas	0.079	16:47
29 Room Air	0.000	16:47
30 Std. Gas	0.080	16:48
31 Room Air	0.000	16:48
32 Std. Gas	0.079	16:49
33 Room Air	0.000	16:49
34 Std. Gas	0.079	16:49
35 Room Air	0.000	16:50
36 Std. Gas	0.079	16:50
37 Room Air	0.000	16:51
38 Std. Gas	0.079	16:51
39 Room Air	0.000	16:52
40 Std. Gas	0.079	16:52
41 Room Air	0.000	16:52
42 Cal Check Stats:		
43 Average	0.0790	
44 Std Dev	0.0003	
45 Rel Std Dev	0.42%	

*Stability test*  
*Deb Kashner*  
*1/2*  
*P. Kashner*

Operator Signature  
DEB KASHUR

Remarks:

Lot#: 617774 Cyl #: 38 Exp date: 28 May 2011

Form 106-I8000

Stability test  
Deb Kashur

2/2

P.17 of 17 run